

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS080522\
 Data File : PS020188.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Aug 2022 15:15
 Operator : AJ\MA
 Sample : PB146794BS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB146794BS

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/08/2022
 Supervised By :Ankita Jodhani 08/08/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 08 07:29:26 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080322.M
 Quant Title : 8080.M
 QLast Update : Thu Aug 04 02:05:37 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.182	6.759	1525.9E6	480.9E6	517.471	590.088
Target Compounds							
1) T	Dalapon	3.000	2.518	2198.6E6	588.2E6	454.130m	541.758
2) T	3,5-DICHL...	6.398	5.944	1840.4E6	654.5E6	473.117	518.339m
3) T	4-Nitroph...	7.013	6.381	545.0E6	184.8E6	392.387	376.209
5) T	DICAMBA	7.370	6.922	5481.9E6	1722.4E6	476.329	483.273
6) T	MCP	7.538	7.018	468.9E6	135.5E6	50.551	49.691
7) T	MCPA	7.686	7.227	630.8E6	210.4E6	48.971	49.988
8) T	DICHLORPROP	8.061	7.541	1371.9E6	505.9E6	479.262	541.039
9) T	2,4-D	8.292	7.833	1647.1E6	657.6E6	546.892	560.638
10) T	Pentachlo...	8.596	8.256	20741.0E6	6925.6E6	530.225	541.224
11) T	2,4,5-TP ...	9.146	8.617	8794.9E6	2959.1E6	516.189	553.241
12) T	2,4,5-T	9.432	8.995	8468.5E6	2609.8E6	466.554	512.690
13) T	2,4-DB	9.983	9.502	1307.8E6	586.6E6	358.351	862.520 #
14) T	DINOSEB	11.134	9.822	7670.0E6	1617.9E6	482.108	426.397
15) T	Picloram	10.960	10.802	16100.2E6	3370.8E6	509.203	446.786
16) T	DCPA	11.404	10.750	11621.2E6	3431.9E6	472.699	517.599

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS080522\
Data File : PS020188.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Aug 2022 15:15
Operator : AJ\MA
Sample : PB146794BS
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PB146794BS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/08/2022
Supervised By :Ankita Jodhani 08/08/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 08 07:29:26 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080322.M
Quant Title : 8080.M
QLast Update : Thu Aug 04 02:05:37 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

