

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS080522\
 Data File : PS020218.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Aug 2022 04:41
 Operator : AJ\MA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 08 07:41:30 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080322.M
 Quant Title : 8080.M
 QLast Update : Thu Aug 04 02:05:37 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.179	6.758	2226.1E6	654.9E6	754.933	803.656
Target Compounds							
1) T	Dalapon	3.006	2.521	3217.1E6	725.4E6	664.512	668.204
2) T	3,5-DICHL...	6.396	5.944	2704.4E6	940.4E6	695.232	744.813
3) T	4-Nitroph...	7.010	6.379	1081.5E6	374.9E6	778.611	763.431
5) T	DICAMBA	7.368	6.921	8168.6E6	2678.5E6	709.779	751.503
6) T	MCPD	7.538	7.018	624.5E6	205.4E6	67.326	75.300
7) T	MCPA	7.686	7.230	886.9E6	301.7E6	68.856	71.666
8) T	DICHLORPROP	8.059	7.540	2005.9E6	694.7E6	700.764	742.871
9) T	2,4-D	8.289	7.832	2334.4E6	878.2E6	775.127	748.733
10) T	Pentachlo...	8.594	8.256	28371.1E6	9635.2E6	725.284	752.982
11) T	2,4,5-TP ...	9.144	8.616	12200.4E6	4076.5E6	716.066	762.149
12) T	2,4,5-T	9.428	8.992	12902.3E6	3969.9E6	710.822	779.896
13) T	2,4-DB	9.979	9.496	2789.2E6	525.5E6	764.264	772.591
14) T	DINOSEB	11.131	9.820	11509.7E6	2892.1E6	723.460	762.239
15) T	Picloram	10.955	10.799	22255.6E6	5322.4E6	703.878	705.454
16) T	DCPA	11.399	10.748	18451.1E6	5070.4E6	750.509	764.723

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS080522\
Data File : PS020218.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Aug 2022 04:41
Operator : AJ\MA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 08 07:41:30 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080322.M
Quant Title : 8080.M
QLast Update : Thu Aug 04 02:05:37 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

