

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS080524\
 Data File : PS027303.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Aug 2024 20:58
 Operator : AR\AJ
 Sample : HSTDICC750
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 06 07:59:08 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080524.M
 Quant Title : 8080.M
 QLast Update : Tue Aug 06 07:56:02 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.186	7.745	1795.0E6	1326.5E6	750.000	750.000
Target Compounds							
1) T	Dalapon	2.605	2.695	2152.1E6	2181.0E6	682.500	682.500
2) T	3,5-DICHL...	6.364	6.699	2661.8E6	1830.1E6	697.500	697.500
3) T	4-Nitroph...	6.984	7.273	995.0E6	823.0E6	682.500	682.500
5) T	DICAMBA	7.368	7.944	7532.4E6	5883.7E6	705.000	705.000
6) T	MCPD	7.547	8.046	532.3E6	395.3E6	70.500	70.500
7) T	MCPA	7.693	8.289	664.0E6	518.2E6	69.750	69.750
8) T	DICHLORPROP	8.063	8.657	1777.6E6	1402.8E6	705.000	705.000
9) T	2,4-D	8.289	8.986	1923.7E6	1571.5E6	705.000	705.000
10) T	Pentachlo...	8.580	9.513	27874.1E6	22115.2E6	712.500	712.500
11) T	2,4,5-TP ...	9.153	9.887	10437.5E6	8826.0E6	712.500	712.500
12) T	2,4,5-T	9.442	10.306	10110.2E6	8719.2E6	712.500	712.500
13) T	2,4-DB	10.011	10.871	1415.4E6	1089.4E6	712.500	712.500
14) T	DINOSEB	11.202	11.249	7685.5E6	5913.3E6	705.000	705.000
15) T	Picloram	11.017	12.341	13428.2E6	12719.7E6	712.500	712.500
16) T	DCPA	11.498	12.289	12663.7E6	9747.6E6	720.000	720.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS080524\
Data File : PS027303.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Aug 2024 20:58
Operator : AR\AJ
Sample : HSTDICC750
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 06 07:59:08 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080524.M
Quant Title : 8080.M
QLast Update : Tue Aug 06 07:56:02 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

