

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS080618\
 Data File : PS001404.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Aug 2018 22:40
 Operator : UA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 07 00:54:03 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS072118.M
 Quant Title : 8080.M
 QLast Update : Sat Jul 21 01:18:58 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.635	11.036	134.7E6	130.3E6	725.828	716.494
Target Compounds							
1) T	Dalapon	3.030	2.427	153.3E6	149.2E6	646.422m	654.701
2) T	3,5-DICHL...	10.051	9.183	184.6E6	168.0E6	683.946	659.476
3) T	4-Nitroph...	11.228	10.182	74512296	64663107	647.435	634.331
5) T	DICAMBA	11.950	11.323	504.6E6	464.3E6	701.782	677.428
6) T	MCPD	12.288	11.547	35630695	29846869	70.892	68.785m
7) T	MCPA	12.539	11.923	53570964	46388215	67.929	71.340m
8) T	DICHLORPROP	13.142	12.447	126.7E6	115.8E6	693.989	653.518
9) T	2,4-D	13.500	12.918	169.1E6	156.3E6	688.892	658.881m
10) T	Pentachlo...	13.889	13.474	1993.8E6	1907.0E6	708.139	692.887m
11) T	2,4,5-TP ...	14.758	14.087	789.6E6	746.9E6	703.580	683.109m
12) T	2,4,5-T	15.155	14.658	778.0E6	577.2E6	699.217	624.807
13) T	2,4-DB	15.906	15.304	100.1E6	88860779	695.457	667.311
14) T	DINOSEB	17.405	15.735	539.7E6	487.0E6	702.721	685.173
15) T	Picloram	17.164	16.917	975.1E6	915.4E6	616.914	664.022
16) T	DCPA	17.744	16.973	960.3E6	887.4E6	711.292	614.442

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS080618\
Data File : PS001404.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Aug 2018 22:40
Operator : UA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
Client Sample ID :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 07 00:54:03 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS072118.M
Quant Title : 8080.M
QLast Update : Sat Jul 21 01:18:58 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

