

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS080620\
 Data File : PS011611.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Aug 2020 07:50
 Operator : DD\AJ
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 06 10:47:13 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS071320.M
 Quant Title : 8080.M
 QLast Update : Mon Jul 13 16:10:29 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds						
5) T	DICAMBA	0.000	7.145	0	3714896	N.D. 1.622
6) T	MCPP	0.000	7.145	0	3714896	N.D. 1.947
15) T	Picloram	10.987	11.014	19458188	5520012	3.084 1.262 #
16) T	DCPA	0.000	11.014	0	5520012	N.D. 1.460



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS080620\
Data File : PS011611.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Aug 2020 07:50
Operator : DD\AJ
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 06 10:47:13 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS071320.M
Quant Title : 8080.M
QLast Update : Mon Jul 13 16:10:29 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA


R.T.: 0.000 min
Exp R.T. : 7.226 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#4 2,4-DCAA

R.T.: 0.000 min
Exp R.T. : 6.918 min
Response: 0
Conc: N.D.




#5 DICAMBA

R.T.: 0.000 min
Exp R.T. : 7.419 min
Response: 0
Conc: N.D.



#5 DICAMBA


R.T.: 7.145 min
Delta R.T.: 0.057 min
Response: 3714896
Conc: 1.62 ng/ml



#6 MCPP


R.T.: 0.000 min
Exp R.T.: 7.594 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#6 MCPP

R.T.: 7.145 min
Delta R.T.: -0.047 min
Response: 3714896
Conc: 1.95 ug/ml




#15 Picloram

R.T.: 10.987 min
Delta R.T.: -0.031 min
Response: 19458188
Conc: 3.08 ng/ml



#15 Picloram


R.T.: 11.014 min
Delta R.T.: 0.015 min
Response: 5520012
Conc: 1.26 ng/ml



#16 DCPA

R.T.: 0.000 min
Exp R.T.: 11.466 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :



#16 DCPA

R.T.: 11.014 min
Delta R.T.: 0.069 min
Response: 5520012
Conc: 1.46 ng/ml