

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS080823\
 Data File : PS023987.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Aug 2023 19:34
 Operator : AR\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 09 00:06:47 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080823.M
 Quant Title : 8080.M
 QLast Update : Tue Aug 08 13:26:17 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.367	6.799	976.8E6	335.9E6	493.212	499.320
Target Compounds							
1) T	Dalapon	2.517	2.582	9041886	6636568	4.121	6.069 #
3) T	4-Nitroph...	6.209	6.392	7164532	766787	6.473	1.937 #
5) T	DICAMBA	6.498	7.007f	1716832	1739696	<MDL	<MDL #
6) T	MCP	6.689	7.007f	1166714	1739696	<MDL	<MDL #
7) T	MCPA	6.852f	0.000	579499	0	<MDL	N.D. #
8) T	DICHLORPROP	7.051f	7.592	6966005	707369	3.096	<MDL #
11) T	2,4,5-TP ...	8.049	8.648	834655	244008	<MDL	<MDL #
12) T	2,4,5-T	8.305	9.064f	1810416	922189	<MDL	<MDL #
14) T	DINOSEB	9.864	0.000	841811	0	<MDL	N.D. #
15) T	Picloram	9.693	10.810	8008353	4252482	<MDL	<MDL #
16) T	DCPA	10.124	10.810	903059	4252482	<MDL	<MDL #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS080823\
Data File : PS023987.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Aug 2023 19:34
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 09 00:06:47 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080823.M
Quant Title : 8080.M
QLast Update : Tue Aug 08 13:26:17 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 2.517 min
Delta R.T.: 0.037 min
Response: 9041886
Conc: 4.12 ng/ml

Instrument :
ECD_S
ClientSampleId :
I.BLK




#1 Dalapon

R.T.: 2.582 min
Delta R.T.: 0.035 min
Response: 6636568
Conc: 6.07 ng/ml




#3 4-Nitrophenol

R.T.: 6.209 min
Delta R.T.: 0.003 min
Response: 7164532
Conc: 6.47 ng/ml



#3 4-Nitrophenol


R.T.: 6.392 min
Delta R.T.: -0.018 min
Response: 766787
Conc: 1.94 ng/ml



#4 2,4-DCAA


R.T.: 6.367 min
Delta R.T.: -0.002 min
Response: 976838126
Conc: 493.21 ng/ml

Instrument :
ECD_S
ClientSampleId :
I.BLK




#4 2,4-DCAA

R.T.: 6.799 min
Delta R.T.: 0.000 min
Response: 335856316
Conc: 499.32 ng/ml



#8 DICHLORPROP

R.T.: 7.051 min
Delta R.T.: -0.050 min
Response: 6966005
Conc: 3.10 ng/ml



#8 DICHLORPROP

R.T.: 7.592 min
Delta R.T.: 0.013 min
Response: 707369
Conc: N.D.