

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS080823\
 Data File : PS023969.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Aug 2023 09:46
 Operator : AR\AJ
 Sample : HSTDICC200
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC200

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 08/09/2023
 Supervised By : Ankita Jodhani 08/09/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 08 13:21:41 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080823.M
 Quant Title : 8080.M
 QLast Update : Tue Aug 08 13:20:23 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.368	6.799	464.8E6	150.6E6	244.824	231.277
Target Compounds							
1) T	Dalapon	2.480	2.547	441.9E6	220.2E6	206.543	207.532m
2) T	3,5-DICHL...	5.741	5.981	653.7E6	213.0E6	223.679	208.364
3) T	4-Nitroph...	6.205	6.411	224.0E6	80161388	209.624	209.642
5) T	DICAMBA	6.510	6.963	1866.5E6	586.5E6	217.490	202.822
6) T	MCP	6.659	7.056	109.6E6	35326167	17.546	18.329
7) T	MCPA	6.781	7.263	152.4E6	55451186	18.269	19.272
8) T	DICHLORPROP	7.100	7.580	488.5E6	152.1E6	225.157	214.312
9) T	2,4-D	7.297	7.864	517.2E6	171.2E6	217.221	205.825
10) T	Pentachlo...	7.526	8.290	7155.5E6	2270.1E6	230.488	214.254
11) T	2,4,5-TP ...	8.048	8.649	2830.9E6	833.0E6	221.754	205.249
12) T	2,4,5-T	8.304	9.021	2301.3E6	738.9E6	208.384	197.029
13) T	2,4-DB	8.809	9.525	345.8E6	90004806	198.824	202.219
14) T	DINOSEB	9.865	9.853	1964.6E6	609.5E6	217.947	207.559
15) T	Picloram	9.690	10.808	2475.0E6	916.9E6	194.779	185.474
16) T	DCPA	10.130	10.777	3533.6E6	1091.3E6	231.222	215.740

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS080823\
Data File : PS023969.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Aug 2023 09:46
Operator : AR\AJ
Sample : HSTDICC200
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC200

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 08/09/2023
Supervised By : Ankita Jodhani 08/09/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 08 13:21:41 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080823.M
Quant Title : 8080.M
QLast Update : Tue Aug 08 13:20:23 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

