

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS080823\
 Data File : PS023970.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Aug 2023 10:10
 Operator : AR\AJ
 Sample : HSTDICC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 08 13:22:03 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080823.M
 Quant Title : 8080.M
 QLast Update : Tue Aug 08 13:20:23 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.368	6.799	986.2E6	332.1E6	519.386	510.067
Target Compounds							
1) T	Dalapon	2.480	2.549	987.6E6	490.2E6	461.579	462.005
2) T	3,5-DICHL...	5.741	5.980	1408.8E6	481.7E6	482.023	471.138
3) T	4-Nitroph...	6.205	6.411	493.8E6	177.3E6	462.159	463.700
5) T	DICAMBA	6.510	6.963	4137.5E6	1359.2E6	482.114	470.014
6) T	MCP	6.661	7.058	279.6E6	87711637	44.772	45.510
7) T	MCPA	6.784	7.265	377.9E6	132.8E6	45.308	46.148
8) T	DICHLORPROP	7.100	7.579	1049.6E6	338.9E6	483.790	477.382
9) T	2,4-D	7.297	7.864	1139.0E6	392.9E6	478.430	472.384
10) T	Pentachlo...	7.527	8.291	15527.9E6	5169.5E6	500.171	487.905
11) T	2,4,5-TP ...	8.049	8.649	6252.7E6	1942.6E6	489.803	478.643
12) T	2,4,5-T	8.304	9.021	5297.3E6	1771.6E6	479.677	472.425
13) T	2,4-DB	8.809	9.525	825.3E6	210.1E6	474.542	471.968
14) T	DINOSEB	9.865	9.853	4333.4E6	1395.5E6	480.728	475.245
15) T	Picloram	9.690	10.809	5972.9E6	2306.3E6	470.053	466.523
16) T	DCPA	10.131	10.776	7645.0E6	2476.4E6	500.250	489.578

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS080823\
Data File : PS023970.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Aug 2023 10:10
Operator : AR\AJ
Sample : HSTDICC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 08 13:22:03 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080823.M
Quant Title : 8080.M
QLast Update : Tue Aug 08 13:20:23 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

