

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS080823\
 Data File : PS023974.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Aug 2023 11:52
 Operator : AR\AJ
 Sample : HSTDICV750
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 ICVPS080823

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 08 13:27:31 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080823.M
 Quant Title : 8080.M
 QLast Update : Tue Aug 08 13:26:17 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.368	6.800	1438.2E6	492.9E6	726.157	732.781
Target Compounds							
1) T	Dalapon	2.480	2.548	1455.5E6	723.5E6	663.321	661.627
2) T	3,5-DICHL...	5.741	5.980	2061.1E6	719.7E6	679.956	686.848
3) T	4-Nitroph...	6.205	6.412	737.9E6	263.9E6	666.704	666.777
5) T	DICAMBA	6.511	6.963	6100.8E6	2052.9E6	693.525	699.996
6) T	MCPD	6.663	7.059	444.6E6	137.3E6	71.289	70.959
7) T	MCPA	6.786	7.267	585.3E6	202.5E6	69.674	69.236
8) T	DICHLORPROP	7.100	7.579	1539.1E6	503.1E6	684.022	688.192
9) T	2,4-D	7.297	7.864	1688.8E6	590.1E6	681.515	692.685
10) T	Pentachlo...	7.527	8.291	22226.3E6	7592.9E6	709.015	706.661
11) T	2,4,5-TP ...	8.049	8.650	9152.6E6	2901.9E6	699.880	705.802
12) T	2,4,5-T	8.304	9.021	7916.7E6	2685.9E6	706.598	710.781
13) T	2,4-DB	8.809	9.525	1251.9E6	318.3E6	703.486	698.218
14) T	DINOSEB	9.866	9.854	6396.0E6	2079.0E6	689.986	693.211
15) T	Picloram	9.689	10.808	9186.2E6	3583.2E6	710.487	719.103
16) T	DCPA	10.131	10.777	11066.9E6	3645.3E6	703.460	709.073

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS080823\
Data File : PS023974.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Aug 2023 11:52
Operator : AR\AJ
Sample : HSTDICV750
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
ICVPS080823

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 08 13:27:31 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080823.M
Quant Title : 8080.M
QLast Update : Tue Aug 08 13:26:17 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

