

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS080823\
 Data File : PS023982.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Aug 2023 17:08
 Operator : AR\AJ
 Sample : PB154665BS
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB154665BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 09 00:02:50 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080823.M
 Quant Title : 8080.M
 QLast Update : Tue Aug 08 13:26:17 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.368	6.800	1021.4E6	345.2E6	515.700	513.231
Target Compounds							
1) T	Dalapon	2.481	2.548	1014.9E6	500.5E6	462.511	457.702
2) T	3,5-DICHL...	5.741	5.980	1464.7E6	500.6E6	483.214	477.804
3) T	4-Nitroph...	6.205	6.411	519.9E6	186.0E6	469.694	469.901
5) T	DICAMBA	6.510	6.963	4281.2E6	1414.7E6	486.672	482.386
6) T	MCP	6.661	7.058	293.6E6	92510234	47.077	47.819
7) T	MCPA	6.784	7.265	394.1E6	138.4E6	46.913	47.308
8) T	DICHLORPROP	7.100	7.580	1087.9E6	351.3E6	483.496	480.473
9) T	2,4-D	7.297	7.865	1188.4E6	409.6E6	479.584	480.813
10) T	Pentachlo...	7.527	8.291	16068.1E6	5373.8E6	512.569	500.126
11) T	2,4,5-TP ...	8.049	8.649	6494.1E6	2021.5E6	496.593	491.664
12) T	2,4,5-T	8.304	9.020	5555.6E6	1852.3E6	495.859	490.175
13) T	2,4-DB	8.809	9.524	874.8E6	220.4E6	491.596	483.573
14) T	DINOSEB	9.865	9.853	4498.9E6	1443.8E6	485.330	481.416
15) T	Picloram	9.689	10.808	6112.9E6	2322.7E6	472.789	466.124
16) T	DCPA	10.130	10.777	8004.1E6	2601.9E6	508.774	506.115

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS080823\
Data File : PS023982.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Aug 2023 17:08
Operator : AR\AJ
Sample : PB154665BS
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PB154665BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 09 00:02:50 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080823.M
Quant Title : 8080.M
QLast Update : Tue Aug 08 13:26:17 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

