

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS080824\
 Data File : PS027353.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Aug 2024 21:58
 Operator : AR\AJ
 Sample : P3506-01
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 RT4534

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 08/09/2024
 Supervised By :Ankita Jodhani 08/09/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 09 05:17:16 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080524.M
 Quant Title : 8080.M
 QLast Update : Tue Aug 06 08:10:22 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
4) S 2,4-DCAA	7.179	7.743	904.6E6	555.5E6	359.650m	298.706

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS080824\
Data File : PS027353.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Aug 2024 21:58
Operator : AR\AJ
Sample : P3506-01
Misc :
ALS Vial : 22 Sample Multiplier: 1



Instrument :
ECD_S
ClientSampleId :
RT4534


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 08/09/2024
Supervised By : Ankita Jodhani 08/09/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 09 05:17:16 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080524.M
Quant Title : 8080.M
QLast Update : Tue Aug 06 08:10:22 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA

R.T.: 7.179 min
Delta R.T.: -0.006 min
Response: 904562702
Conc: 359.65 ng/ml

Instrument :
ECD_S
ClientSampleId :
RT4534

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/09/2024
Supervised By :Ankita Jodhani 08/09/2024



#4 2,4-DCAA

R.T.: 7.743 min
Delta R.T.: -0.002 min
Response: 555524590
Conc: 298.71 ng/ml