

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS080919\
 Data File : PS005982.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Aug 2019 17:03
 Operator : SM\AJ
 Sample : K4152-31
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 2S-E1-E4-COMP-(0-1)

Manual Integrations
 APPROVED

Ankita
 8/12/2019 10:18:37 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 10 03:12:56 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS072619.M
 Quant Title : 8080.M
 QLast Update : Sat Jul 27 00:42:55 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
4) S 2,4-DCAA	5.573	5.293	629.4E6	441.1E6	433.791	443.448m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS080919\
Data File : PS005982.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Aug 2019 17:03
Operator : SM\AJ
Sample : K4152-31
Misc :
ALS Vial : 17 Sample Multiplier: 1



Instrument :
ECD_S
ClientSampleId :
2S-E1-E4-COMP-(0-1)


Manual Integrations
APPROVED

Ankita
8/12/2019 10:18:37 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 10 03:12:56 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS072619.M
Quant Title : 8080.M
QLast Update : Sat Jul 27 00:42:55 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA

R.T.: 5.573 min
Delta R.T.: 0.017 min
Response: 629351054
Conc: 433.79 ng/ml

Instrument :
ECD_S
ClientSampleId :
2S-E1-E4-COMP-(0-1)

Manual Integrations
APPROVED

Ankita
8/12/2019 10:18:37 AM



#4 2,4-DCAA

R.T.: 5.293 min
Delta R.T.: 0.017 min
Response: 441088229
Conc: 443.45 ng/ml m