

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS080922\
 Data File : PS020246.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Aug 2022 20:05
 Operator : AJ\MA
 Sample : N4091-09MSD
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 RO-124019MSD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/10/2022
 Supervised By :Ankita Jodhani 08/10/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 10 01:37:07 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080322.M
 Quant Title : 8080.M
 QLast Update : Thu Aug 04 02:05:37 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.177	6.757	702.8E6	325.9E6	238.351	399.957 #
Target Compounds							
1) T	Dalapon	3.000	2.518	917.6E6	257.6E6	189.529	237.299m#
2) T	3,5-DICHL...	6.395	5.944	852.9E6	253.7E6	219.252	200.937
3) T	4-Nitroph...	7.007	6.378	434.5E6	93622998	312.830	190.631m#
5) T	DICAMBA	7.367	6.920	2649.0E6	789.0E6	230.178	221.383m
6) T	MCP	7.533	7.014	183.5E6	53665989	19.780m	19.673
7) T	MCPA	7.680	7.222	289.5E6	99059101	22.473	23.531m
8) T	DICHLORPROP	8.057	7.540	565.9E6	239.0E6	197.678	255.627 #
9) T	2,4-D	8.285	7.830	785.0E6	319.4E6	260.664	272.320
10) T	Pentachlo...	8.593	8.255	9119.5E6	2898.3E6	233.132	226.500
11) T	2,4,5-TP ...	9.143	8.616	3966.9E6	1268.3E6	232.827	237.131
12) T	2,4,5-T	9.430	8.994	3797.3E6	1171.3E6	209.203	230.106
13) T	2,4-DB	9.979	9.498	577.6E6	320.5E6	158.273	471.219 #
14) T	DINOSEB	11.117	9.822	355.5E6	362.0E6	22.348m	95.395 #
15) T	Picloram	10.955	10.800	7181.4E6	2060.6E6	227.126	273.117
16) T	DCPA	11.401	10.748	5359.4E6	1743.9E6	217.996	263.011

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS080922\
Data File : PS020246.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Aug 2022 20:05
Operator : AJ\MA
Sample : N4091-09MSD
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
RO-124019MSD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/10/2022
Supervised By :Ankita Jodhani 08/10/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 10 01:37:07 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080322.M
Quant Title : 8080.M
QLast Update : Thu Aug 04 02:05:37 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

