

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS080923\
 Data File : PS024003.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Aug 2023 14:07
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 09 21:56:32 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080823.M
 Quant Title : 8080.M
 QLast Update : Tue Aug 08 13:26:17 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.368	6.799	1512.3E6	518.1E6	763.555	770.333
Target Compounds							
1) T	Dalapon	2.482	2.548	1444.2E6	717.3E6	658.196	656.020
2) T	3,5-DICHL...	5.741	5.980	2164.9E6	759.3E6	714.223	724.652
3) T	4-Nitroph...	6.204	6.411	802.0E6	280.7E6	724.575	709.099
5) T	DICAMBA	6.510	6.962	6407.0E6	2156.0E6	728.330	735.145
6) T	MCPD	6.663	7.059	465.0E6	142.0E6	74.563	73.384
7) T	MCPA	6.785	7.267	624.5E6	214.1E6	74.345	73.195
8) T	DICHLORPROP	7.099	7.578	1618.2E6	528.9E6	719.142	723.439
9) T	2,4-D	7.296	7.863	1802.7E6	624.5E6	727.479	733.046
10) T	Pentachlo...	7.526	8.290	23331.2E6	7934.3E6	744.262	738.431
11) T	2,4,5-TP ...	8.048	8.648	9681.6E6	3052.0E6	740.334	742.298
12) T	2,4,5-T	8.303	9.019	8500.0E6	2844.7E6	758.653	752.803
13) T	2,4-DB	8.808	9.523	1355.6E6	341.4E6	761.732	748.928
14) T	DINOSEB	9.864	9.851	6750.9E6	2157.5E6	728.271	719.404
15) T	Picloram	9.688	10.807	9778.3E6	3694.0E6	756.283	741.337
16) T	DCPA	10.130	10.775	11697.8E6	3825.3E6	743.560	744.073

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS080923\
Data File : PS024003.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Aug 2023 14:07
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 09 21:56:32 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080823.M
Quant Title : 8080.M
QLast Update : Tue Aug 08 13:26:17 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

