

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS081120\
 Data File : PS011714.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Aug 2020 15:29
 Operator : DD\AJ
 Sample : HSTDICV750
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 ICVPS081120

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 11 16:12:28 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081120.M
 Quant Title : 8080.M
 QLast Update : Tue Aug 11 15:43:16 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.220	6.915	823.4E6	402.3E6	754.924	752.291
Target Compounds							
1) T	Dalapon	3.034	2.621	858.5E6	524.5E6	689.222	693.973
2) T	3,5-DICHL...	6.434	6.067	1016.0E6	592.5E6	700.800	720.250
3) T	4-Nitroph...	7.051	6.522	382.1E6	236.8E6	648.626	649.452
5) T	DICAMBA	7.411	7.084	2941.4E6	1546.2E6	718.115	715.687
6) T	MCPD	7.581	7.180	230.9E6	126.3E6	72.271	71.481
7) T	MCPA	7.730	7.396	324.8E6	190.8E6	71.407	69.628
8) T	DICHLORPROP	8.103	7.709	715.4E6	400.4E6	712.036	709.096
9) T	2,4-D	8.332	8.004	802.6E6	518.2E6	710.845	705.403
10) T	Pentachlo...	8.643	8.433	10385.8E6	4817.2E6	722.160	708.953
11) T	2,4,5-TP ...	9.191	8.793	4211.4E6	2107.3E6	724.260	720.782
12) T	2,4,5-T	9.477	9.173	4269.3E6	2139.4E6	720.462	722.743
13) T	2,4-DB	10.030	9.679	575.0E6	300.7E6	723.653	719.065
14) T	DINOSEB	11.189	10.012	3358.4E6	1493.5E6	715.507	702.531
15) T	Picloram	11.012	10.996	5812.8E6	3324.9E6	743.736	733.575
16) T	DCPA	11.454	10.940	5599.9E6	2662.7E6	736.502	724.163

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS081120\
Data File : PS011714.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Aug 2020 15:29
Operator : DD\AJ
Sample : HSTDICV750
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
ICVPS081120

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 11 16:12:28 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081120.M
Quant Title : 8080.M
QLast Update : Tue Aug 11 15:43:16 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

