

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS081224\
 Data File : PS027379.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Aug 2024 19:36
 Operator : AR\AJ
 Sample : P3520-01MS
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 OU4-CF-08-080624MS

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 08/13/2024
 Supervised By : Ankita Jodhani 08/14/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 13 04:11:04 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080524.M
 Quant Title : 8080.M
 QLast Update : Tue Aug 06 08:10:22 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.178	7.742	359.1E6	167.5E6	142.761	90.091 #
Target Compounds							
1) T	Dalapon	2.607	2.696	1069.1E6	1284.6E6	327.155m	384.980m
2) T	3,5-DICHL...	6.358	6.697	1117.6E6	713.9E6	280.690	260.508
5) T	DICAMBA	7.360	7.941	3908.0E6	3133.0E6	354.095	363.732
6) T	MCPPE	7.537	8.041	364.5E6	245.1E6	47.904m	43.039
7) T	MCPA	7.683	8.284	496.4E6	415.0E6	51.242m	54.938m
8) T	DICHLORPROP	8.053	8.653	1268.8E6	957.3E6	479.523	457.953
9) T	2,4-D	8.279	8.982	1733.9E6	1590.2E6	608.155	682.589
10) T	Pentachlo...	8.571	9.509	5608.3E6	4321.1E6	141.878	135.766
11) T	2,4,5-TP ...	9.142	9.895	7455.5E6	15647.8E6	491.571m	1225.249 #
12) T	2,4,5-T	9.430	10.302	7769.9E6	6248.5E6	531.978	495.551
13) T	2,4-DB	9.999	10.866	808.9E6	668.7E6	394.443	423.052
14) T	DINOSEB	11.188	11.245	1077.8E6	941.0E6	96.707	110.480
15) T	Picloram	11.003	12.336	6359.5E6	5679.8E6	329.844	314.827
16) T	DCPA	11.485	12.284	9625.5E6	7014.2E6	532.831	504.389m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS081224\
Data File : PS027379.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Aug 2024 19:36
Operator : AR\AJ
Sample : P3520-01MS
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
OU4-CF-08-080624MS

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 08/13/2024
Supervised By : Ankita Jodhani 08/14/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 13 04:11:04 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080524.M
Quant Title : 8080.M
QLast Update : Tue Aug 06 08:10:22 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

