

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS081224\
 Data File : PS027405.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Aug 2024 12:29
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 08/14/2024
 Supervised By : Ankita Jodhani 08/14/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 14 03:15:38 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080524.M
 Quant Title : 8080.M
 QLast Update : Tue Aug 06 08:10:22 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.175	7.741	1872.2E6	1366.2E6	744.367	734.600
Target Compounds							
1) T	Dalapon	2.601	2.695	2140.0E6	2203.4E6	654.845	660.343
2) T	3,5-DICHL...	6.356	6.696	2758.7E6	1879.0E6	692.846	685.618
3) T	4-Nitroph...	6.974	7.269	1017.1E6	839.5E6	672.748	670.964m
5) T	DICAMBA	7.357	7.939	7722.2E6	6061.2E6	699.698	703.675
6) T	MCP	7.536	8.041	545.4E6	404.1E6	71.666	70.971
7) T	MCPA	7.683	8.284	692.5E6	538.6E6	71.483	71.293m
8) T	DICHLORPROP	8.051	8.652	1848.6E6	1443.4E6	698.686	690.464
9) T	2,4-D	8.277	8.980	1990.8E6	1619.2E6	698.253	695.041
10) T	Pentachlo...	8.568	9.508	28867.1E6	22860.3E6	730.274	718.265
11) T	2,4,5-TP ...	9.140	9.881	10995.4E6	9099.5E6	724.971	712.511
12) T	2,4,5-T	9.427	10.300	10664.4E6	9140.0E6	730.152	724.864
13) T	2,4-DB	9.995	10.865	1499.7E6	1122.2E6	731.280	709.948
14) T	DINOSEB	11.183	11.243	8090.2E6	5958.0E6	725.888	699.527
15) T	Picloram	11.000	12.335	13855.7E6	12902.2E6	718.642	715.155
16) T	DCPA	11.481	12.283	13424.6E6	10008.3E6	743.137	719.692

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS081224\
Data File : PS027405.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Aug 2024 12:29
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 08/14/2024
Supervised By : Ankita Jodhani 08/14/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 14 03:15:38 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080524.M
Quant Title : 8080.M
QLast Update : Tue Aug 06 08:10:22 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

