

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS081318\
 Data File : PS001463.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Aug 2018 20:26
 Operator : UA
 Sample : J4405-04
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 14 01:08:14 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS072118.M
 Quant Title : 8080.M
 QLast Update : Sat Jul 21 01:18:58 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.613	11.033	104.1E6	98401985	560.770	541.042
Target Compounds							
1) T	Dalapon	3.010	2.401	1213762	3809860	5.119	16.719 #
2) T	3,5-DICHL...	10.037	0.000	2400551	0	8.894	N.D. #
3) T	4-Nitroph...	0.000	10.191	0	39178774	N.D.	384.335 #
5) T	DICAMBA	11.928	11.277	2724377	2817453	3.789	4.111
6) T	MCPD	12.305	11.579	13493100	2737446	26.846	6.309 #
7) T	MCPA	12.575	11.958	4097853	6206847	5.196	9.545 #
10) T	Pentachlo...	13.868	13.471	12445281	11194422	4.420	4.067
11) T	2,4,5-TP ...	14.736	14.085	4297994	4292379	3.830	3.926
12) T	2,4,5-T	15.137	14.656	4901049	5040922	4.405	5.457
14) T	DINOSEB	17.383	15.733	3070441	2386061	3.998	3.357
15) T	Picloram	17.147	16.913	9316919	16793869	5.894	12.183 #
16) T	DCPA	17.734	16.970	10342354	18812745	7.661	13.026 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS081318\
Data File : PS001463.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Aug 2018 20:26
Operator : UA
Sample : J4405-04
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 14 01:08:14 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS072118.M
Quant Title : 8080.M
QLast Update : Sat Jul 21 01:18:58 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 3.010 min
Delta R.T.: -0.003 min
Response: 1213762
Conc: 5.12 ng/ml

Instrument :
ECD_S
ClientSampleId :




#1 Dalapon

R.T.: 2.401 min
Delta R.T.: -0.024 min
Response: 3809860
Conc: 16.72 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 10.037 min
Delta R.T.: 0.001 min
Response: 2400551
Conc: 8.89 ng/ml



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 0.000 min
Exp R.T. : 9.186 min
Response: 0
Conc: N.D.



#3 4-Nitrophenol


R.T.: 0.000 min
Exp R.T.: 11.213 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#3 4-Nitrophenol

R.T.: 10.191 min
Delta R.T.: 0.006 min
Response: 39178774
Conc: 384.34 ng/ml




#4 2,4-DCAA

R.T.: 11.613 min
Delta R.T.: -0.007 min
Response: 104050752
Conc: 560.77 ng/ml



#4 2,4-DCAA


R.T.: 11.033 min
Delta R.T.: -0.005 min
Response: 98401985
Conc: 541.04 ng/ml



#5 DICAMBA


R.T.: 11.928 min
Delta R.T.: -0.005 min
Response: 2724377
Conc: 3.79 ng/ml

Instrument :
ECD_S
ClientSampleId :




#5 DICAMBA

R.T.: 11.277 min
Delta R.T.: -0.050 min
Response: 2817453
Conc: 4.11 ng/ml




#6 MCP

R.T.: 12.305 min
Delta R.T.: 0.035 min
Response: 13493100
Conc: 26.85 ug/ml



#6 MCP


R.T.: 11.579 min
Delta R.T.: 0.033 min
Response: 2737446
Conc: 6.31 ug/ml



#7 MCPA


R.T.: 12.575 min
Delta R.T.: 0.054 min
Response: 4097853
Conc: 5.20 ug/ml

Instrument :
ECD_S
ClientSampleId :




#7 MCPA

R.T.: 11.958 min
Delta R.T.: 0.036 min
Response: 6206847
Conc: 9.55 ug/ml




#10 Pentachlorophenol

R.T.: 13.868 min
Delta R.T.: -0.006 min
Response: 12445281
Conc: 4.42 ng/ml



#10 Pentachlorophenol


R.T.: 13.471 min
Delta R.T.: -0.007 min
Response: 11194422
Conc: 4.07 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 14.736 min
Delta R.T.: -0.008 min
Response: 4297994
Conc: 3.83 ng/ml

Instrument :
ECD_S
ClientSampleId :




#11 2,4,5-TP (SILVEX)

R.T.: 14.085 min
Delta R.T.: -0.007 min
Response: 4292379
Conc: 3.93 ng/ml




#12 2,4,5-T

R.T.: 15.137 min
Delta R.T.: -0.005 min
Response: 4901049
Conc: 4.40 ng/ml



#12 2,4,5-T


R.T.: 14.656 min
Delta R.T.: -0.006 min
Response: 5040922
Conc: 5.46 ng/ml



#14 DINOSEB


R.T.: 17.383 min
Delta R.T.: -0.009 min
Response: 3070441
Conc: 4.00 ng/ml

Instrument :
ECD_S
ClientSampleId :




#14 DINOSEB

R.T.: 15.733 min
Delta R.T.: -0.007 min
Response: 2386061
Conc: 3.36 ng/ml




#15 Picloram

R.T.: 17.147 min
Delta R.T.: -0.005 min
Response: 9316919
Conc: 5.89 ng/ml



#15 Picloram


R.T.: 16.913 min
Delta R.T.: -0.009 min
Response: 16793869
Conc: 12.18 ng/ml



#16 DCPA

R.T.: 17.734 min
Delta R.T.: 0.002 min
Response: 10342354
Conc: 7.66 ng/ml

Instrument :
ECD_S
ClientSampleId :



#16 DCPA

R.T.: 16.970 min
Delta R.T.: -0.008 min
Response: 18812745
Conc: 13.03 ng/ml