

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS081419\
 Data File : PS006054.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Aug 2019 14:13
 Operator : SM\AJ
 Sample : HSTDICC1500
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC1500

Manual Integrations
 APPROVED

Ankita
 8/22/2019 2:16:05 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 14 15:08:35 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081419.M
 Quant Title : 8080.M
 QLast Update : Wed Aug 14 14:54:21 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.487	5.218	1240.5E6	725.3E6	1422.218	1468.485
Target Compounds							
1) T	Dalapon	2.053	1.762	1500.7E6	902.7E6	1380.407	1458.327
2) T	3,5-DICHL...	4.853	4.520	1462.7E6	832.7E6	1347.435	1369.848
3) T	4-Nitroph...	5.335	4.900	540.7E6	352.6E6	1460.429	1438.594
5) T	DICAMBA	5.641	5.347	4406.2E6	2504.6E6	1362.754	1345.441
6) T	MCPD	5.796	5.435	338.0E6	188.1E6	152.677	155.082
7) T	MCPA	5.930	5.615	469.8E6	280.7E6	151.348	136.824
8) T	DICHLORPROP	6.269	5.888	1050.4E6	617.0E6	1379.021	1382.168
9) T	2,4-D	6.482	6.157	1178.0E6	763.5E6	1454.987	1383.655
10) T	Pentachlo...	6.747	6.527	15158.2E6	8652.2E6	1251.007	1248.163
11) T	2,4,5-TP ...	7.282	6.890	6152.5E6	3527.5E6	1330.878	1339.333
12) T	2,4,5-T	7.553	7.248	6594.2E6	3490.9E6	1312.581	1399.077m
13) T	2,4-DB	8.082	7.735	919.7E6	487.1E6	1490.234	1473.514
14) T	DINOSEB	9.175	8.046	4807.8E6	2628.2E6	1291.199	1358.215
15) T	Picloram	9.003	8.982	10349.9E6	6145.2E6	1382.751	1347.484
16) T	DCPA	9.447	8.943	7760.5E6	4711.4E6	1315.973	1330.523

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS081419\
Data File : PS006054.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Aug 2019 14:13
Operator : SM\AJ
Sample : HSTDICC1500
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC1500

Manual Integrations
APPROVED

Ankita
8/22/2019 2:16:05 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 14 15:08:35 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081419.M
Quant Title : 8080.M
QLast Update : Wed Aug 14 14:54:21 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

