

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS081420\
 Data File : PS011782.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Aug 2020 10:46
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 14 12:14:40 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081120.M
 Quant Title : 8080.M
 QLast Update : Tue Aug 11 15:43:16 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.219	6.915	790.8E6	428.5E6	724.987	801.166
Target Compounds							
1) T	Dalapon	3.035	2.622	860.6E6	539.4E6	690.917	713.690
2) T	3,5-DICHL...	6.434	6.067	1008.7E6	604.4E6	695.803	734.722
3) T	4-Nitroph...	7.051	6.523	356.7E6	245.8E6	605.593	674.105
5) T	DICAMBA	7.410	7.083	2967.2E6	1657.5E6	724.408	767.194
6) T	MCPD	7.580	7.180	234.8E6	137.2E6	73.473	77.600
7) T	MCPA	7.729	7.396	329.9E6	206.3E6	72.530	75.305
8) T	DICHLORPROP	8.101	7.709	717.7E6	430.4E6	714.336	762.204
9) T	2,4-D	8.332	8.004	785.5E6	559.6E6	695.657	761.745
10) T	Pentachlo...	8.641	8.432	10324.0E6	5013.4E6	717.865	737.831
11) T	2,4,5-TP ...	9.190	8.793	4172.1E6	2245.9E6	717.500	768.175
12) T	2,4,5-T	9.476	9.173	4152.4E6	2249.4E6	700.730	759.920
13) T	2,4-DB	10.029	9.679	538.7E6	315.0E6	677.946	753.086
14) T	DINOSEB	11.187	10.011	3222.4E6	1603.1E6	686.549	754.049
15) T	Picloram	11.012	10.997	5152.4E6	3362.5E6	659.245	741.877
16) T	DCPA	11.453	10.938	5206.1E6	2835.4E6	684.703	771.121

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS081420\
Data File : PS011782.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Aug 2020 10:46
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 14 12:14:40 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081120.M
Quant Title : 8080.M
QLast Update : Tue Aug 11 15:43:16 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

