

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS081519\
 Data File : PS006129.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Aug 2019 19:12
 Operator : SM\AJ
 Sample : WATER IDOC 04
 Misc : FOR DHAVAL DAVE
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 WATER IDOC 04

Manual Integrations
 APPROVED

Ankita
 8/16/2019 2:47:37 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 16 03:57:27 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081419.M
 Quant Title : 8080.M
 QLast Update : Wed Aug 14 15:09:52 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.487	5.217	481.4E6	258.9E6	557.743	526.460
Target Compounds							
1) T	Dalapon	2.050	1.759	489.4E6	277.8E6	449.166	442.670
2) T	3,5-DICHL...	4.854	4.519	520.6E6	297.3E6	482.872	490.851
3) T	4-Nitroph...	5.346	4.903	182.4E6	117.3E6	485.819	473.664
5) T	DICAMBA	5.638	5.344	1630.0E6	913.1E6	507.525	495.065
6) T	MCPD	5.787	5.426	94329877	61536752	41.917	49.743
7) T	MCPA	5.920	5.604	155.4E6	104.5E6	49.230	51.126
8) T	DICHLORPROP	6.269	5.886	379.0E6	220.5E6	499.816	495.862
9) T	2,4-D	6.489	6.159	381.9E6	269.3E6	468.653	489.978
10) T	Pentachlo...	6.743	6.523	6083.3E6	3496.1E6	514.622	517.183
11) T	2,4,5-TP ...	7.280	6.887	2312.6E6	1319.4E6	506.940	507.061
12) T	2,4,5-T	7.555	7.247	2564.2E6	1254.4E6	518.589	504.577m
13) T	2,4-DB	8.086	7.734	279.9E6	160.2E6	449.426	481.461
14) T	DINOSEB	9.170	8.042	1922.9E6	953.8E6	525.264	496.541
15) T	Picloram	9.008	8.981	3561.7E6	2163.5E6	478.680	479.623
16) T	DCPA	9.443	8.938	3038.5E6	1806.5E6	524.289	518.055

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS081519\
Data File : PS006129.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2019 19:12
Operator : SM\AJ
Sample : WATER IDOC 04
Misc : FOR DHAVAL DAVE
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleID :
WATER IDOC 04

Manual Integrations
APPROVED

Ankita
8/16/2019 2:47:37 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 16 03:57:27 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081419.M
Quant Title : 8080.M
QLast Update : Wed Aug 14 15:09:52 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

