

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS081522\
 Data File : PS020270.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Aug 2022 10:57
 Operator : AJ\MA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 15 14:08:40 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080322.M
 Quant Title : 8080.M
 QLast Update : Thu Aug 04 02:05:37 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.179	6.757	2170.0E6	666.1E6	735.918	817.391
Target Compounds							
1) T	Dalapon	3.004	2.520	3565.5E6	790.9E6	736.472	728.477
2) T	3,5-DICHL...	6.396	5.944	2653.9E6	958.6E6	682.242	759.240
3) T	4-Nitroph...	7.010	6.379	1053.8E6	385.0E6	758.640	783.909
5) T	DICAMBA	7.367	6.920	8030.9E6	2763.9E6	697.818	775.480
6) T	MCPD	7.537	7.017	637.5E6	211.6E6	68.723	77.573
7) T	MCPA	7.685	7.229	886.3E6	309.1E6	68.808	73.426
8) T	DICHLORPROP	8.058	7.539	1925.4E6	712.4E6	672.644	761.805
9) T	2,4-D	8.288	7.830	2350.2E6	905.0E6	780.373	771.578
10) T	Pentachlo...	8.592	8.253	28072.9E6	9793.2E6	717.661	765.330
11) T	2,4,5-TP ...	9.143	8.614	12083.1E6	4167.8E6	709.181	779.222
12) T	2,4,5-T	9.428	8.991	12862.2E6	4032.3E6	708.609	792.156
13) T	2,4-DB	9.978	9.495	2906.5E6	539.3E6	796.406	792.881
14) T	DINOSEB	11.129	9.819	12167.6E6	2960.4E6	764.810	780.231
15) T	Picloram	10.955	10.798	25044.0E6	5909.1E6	792.067	783.213
16) T	DCPA	11.399	10.746	19073.1E6	5150.8E6	775.810	776.854

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS081522\
Data File : PS020270.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2022 10:57
Operator : AJ\MA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 15 14:08:40 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080322.M
Quant Title : 8080.M
QLast Update : Thu Aug 04 02:05:37 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

