

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS081619\
 Data File : PS006155.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Aug 2019 16:02
 Operator : SM\AJ
 Sample : K4265-04
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 17 01:32:45 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081419.M
 Quant Title : 8080.M
 QLast Update : Wed Aug 14 15:09:52 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.508	5.237	219.8E6	141.3E6	254.648	287.227
Target Compounds							
1) T	Dalapon	2.045	1.760	81087510	83835661	74.418	133.604 #
2) T	3,5-DICHL...	0.000	4.563	0	51544195	N.D.	85.098
9) T	2,4-D	6.566	0.000	216.3E6	0	265.510	N.D. #
15) T	Picloram	9.071	9.018	95380721	65760032	12.819	14.578
16) T	DCPA	0.000	9.018	0	65760032	N.D.	18.858



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS081619\
Data File : PS006155.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Aug 2019 16:02
Operator : SM\AJ
Sample : K4265-04
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 17 01:32:45 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081419.M
Quant Title : 8080.M
QLast Update : Wed Aug 14 15:09:52 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 2.045 min
Delta R.T.: -0.026 min
Response: 81087510
Conc: 74.42 ng/ml

Instrument :
ECD_S
ClientSampleId :




#1 Dalapon

R.T.: 1.760 min
Delta R.T.: -0.019 min
Response: 83835661
Conc: 133.60 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 0.000 min
Exp R.T. : 4.872 min
Response: 0
Conc: N.D.



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 4.563 min
Delta R.T.: 0.025 min
Response: 51544195
Conc: 85.10 ng/ml



#4 2,4-DCAA


R.T.: 5.508 min
Delta R.T.: 0.003 min
Response: 219812309
Conc: 254.65 ng/ml

Instrument :
ECD_S
ClientSampleId :




#4 2,4-DCAA

R.T.: 5.237 min
Delta R.T.: 0.001 min
Response: 141273184
Conc: 287.23 ng/ml




#9 2,4-D

R.T.: 6.566 min
Delta R.T.: 0.063 min
Response: 216333755
Conc: 265.51 ng/ml



#9 2,4-D


R.T.: 0.000 min
Exp R.T.: 6.176 min
Response: 0
Conc: N.D.



#15 Picloram


R.T.: 9.071 min
Delta R.T.: 0.047 min
Response: 95380721
Conc: 12.82 ng/ml

Instrument :
ECD_S
ClientSampleId :




#15 Picloram

R.T.: 9.018 min
Delta R.T.: 0.018 min
Response: 65760032
Conc: 14.58 ng/ml



#16 DCPA

R.T.: 0.000 min
Exp R.T. : 9.465 min
Response: 0
Conc: N.D.



#16 DCPA

R.T.: 9.018 min
Delta R.T.: 0.058 min
Response: 65760032
Conc: 18.86 ng/ml