

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS081619\
 Data File : PS006154.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Aug 2019 15:40
 Operator : SM\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 17 01:32:20 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081419.M
 Quant Title : 8080.M
 QLast Update : Wed Aug 14 15:09:52 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.513	5.243	651.6E6	380.6E6	754.904	773.769
Target Compounds							
1) T	Dalapon	2.077	1.786	719.3E6	424.1E6	660.108	675.897
2) T	3,5-DICHL...	4.879	4.545	793.7E6	443.7E6	736.138	732.593
3) T	4-Nitroph...	5.366	4.928	274.0E6	180.7E6	729.863	729.611
5) T	DICAMBA	5.664	5.371	2303.0E6	1343.1E6	717.068	728.138
6) T	MCPD	5.815	5.454	156.4E6	95489614	69.510	77.188
7) T	MCPA	5.948	5.633	238.6E6	153.5E6	75.579	75.108
8) T	DICHLORPROP	6.295	5.913	552.0E6	324.5E6	727.827	729.849
9) T	2,4-D	6.512	6.184	596.7E6	397.6E6	732.359	723.280
10) T	Pentachlo...	6.769	6.549	8702.0E6	5009.3E6	736.155	741.027
11) T	2,4,5-TP ...	7.305	6.913	3344.7E6	1910.7E6	733.203	734.301
12) T	2,4,5-T	7.579	7.273	3669.2E6	1796.0E6	742.058	722.415
13) T	2,4-DB	8.110	7.760	445.9E6	241.7E6	715.898	726.337
14) T	DINOSEB	9.197	8.069	2706.8E6	1410.3E6	739.390	734.200
15) T	Picloram	9.031	9.007	4841.4E6	2897.3E6	650.667	642.295
16) T	DCPA	9.469	8.964	4324.4E6	2628.2E6	746.155	753.677

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS081619\
Data File : PS006154.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Aug 2019 15:40
Operator : SM\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 17 01:32:20 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081419.M
Quant Title : 8080.M
QLast Update : Wed Aug 14 15:09:52 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

