

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS081619\
 Data File : PS006180.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Aug 2019 01:15
 Operator : SM\AJ
 Sample : K4321-03MSD
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 SP-2MSD

Manual Integrations
 APPROVED

Ankita
 8/20/2019 9:48:27 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 17 02:12:16 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081419.M
 Quant Title : 8080.M
 QLast Update : Wed Aug 14 15:09:52 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.499	5.230	216.3E6	125.7E6	250.600	255.576m
Target Compounds							
1) T	Dalapon	2.060	1.769	315.4E6	370.6E6	289.462	590.617 #
2) T	3,5-DICHL...	4.867	4.533	248.0E6	136.6E6	230.053	225.530m
5) T	DICAMBA	5.650	5.357	739.3E6	418.9E6	230.190	227.115m
6) T	MCP	5.799	5.438	47664496	28716374	21.180	23.213
7) T	MCPA	5.932	5.615	80416703	65873582	25.472	32.232 #
8) T	DICHLORPROP	6.282	5.899	206.3E6	120.2E6	272.077	270.463
9) T	2,4-D	6.502	6.172	244.8E6	168.4E6	300.461	306.407
10) T	Pentachlo...	6.759	6.537	316.2E6	186.7E6	26.746	27.619
11) T	2,4,5-TP ...	7.293	6.900	1168.8E6	683.5E6	256.208	262.685
12) T	2,4,5-T	7.569	7.261	1334.8E6	670.9E6	269.957	269.847
13) T	2,4-DB	8.099	7.747	129.0E6	83039412	207.152m	249.515
14) T	DINOSEB	9.185	8.057	201.9E6	83368076	55.150	43.402
15) T	Picloram	9.023	8.996	1800.4E6	1130.2E6	241.964	250.561
16) T	DCPA	9.456	8.951	1657.7E6	795.5E6	286.024	228.120

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS081619\
Data File : PS006180.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Aug 2019 01:15
Operator : SM\AJ
Sample : K4321-03MSD
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
SP-2MSD

Manual Integrations
APPROVED

Ankita
8/20/2019 9:48:27 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 17 02:12:16 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081419.M
Quant Title : 8080.M
QLast Update : Wed Aug 14 15:09:52 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

