

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS081619\
 Data File : PS006190.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Aug 2019 04:57
 Operator : SM\AJ
 Sample : K3616-07MS
 Misc :
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 JC-03-081519-AMS

Manual Integrations
 APPROVED

Ankita
 8/19/2019 3:30:06 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 17 05:51:43 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081419.M
 Quant Title : 8080.M
 QLast Update : Wed Aug 14 15:09:52 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.493	5.230	49406567	44453502	57.236	90.380 #
Target Compounds							
1) T	Dalapon	2.049	1.759	484.6E6	392.6E6	444.766m	625.648 #
2) T	3,5-DICHL...	4.854	4.521	580.7E6	331.9E6	538.575	547.924
3) T	4-Nitroph...	5.342	4.904	190.5E6	111.1E6	507.414	448.328
5) T	DICAMBA	5.640	5.347	1660.2E6	960.9E6	516.925	520.936
6) T	MCPD	5.790	5.430	123.1E6	70068535	54.686	56.639
7) T	MCPA	5.923	5.608	192.4E6	122.8E6	60.929	60.084
8) T	DICHLORPROP	6.269	5.888	448.0E6	263.1E6	590.740	591.872
9) T	2,4-D	6.485	6.158	571.4E6	358.0E6	701.300	651.290
10) T	Pentachlo...	6.746	6.525	6880.0E6	3944.0E6	582.023	583.441
11) T	2,4,5-TP ...	7.280	6.888	2838.8E6	1543.7E6	622.290	593.271
12) T	2,4,5-T	7.554	7.248	3121.9E6	1553.4E6	631.372	624.862m
13) T	2,4-DB	8.083	7.735	446.7E6	194.6E6	717.263	584.861
14) T	DINOSEB	9.172	8.045	2127.2E6	1088.1E6	581.071	566.491
15) T	Picloram	9.006	8.982	3947.4E6	2360.2E6	530.522	523.218
16) T	DCPA	9.445	8.940	3638.3E6	2102.5E6	627.782	602.921

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS081619\
Data File : PS006190.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Aug 2019 04:57
Operator : SM\AJ
Sample : K3616-07MS
Misc :
ALS Vial : 44 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
JC-03-081519-AMS

Manual Integrations
APPROVED

Ankita
8/19/2019 3:30:06 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 17 05:51:43 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081419.M
Quant Title : 8080.M
QLast Update : Wed Aug 14 15:09:52 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

