

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS081621\
 Data File : PS016428.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Aug 2021 00:33
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 17 06:32:09 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080221.M
 Quant Title : 8080.M
 QLast Update : Mon Aug 02 14:10:19 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.184	6.765	1102.0E6	389.4E6	749.083	746.573
Target Compounds							
1) T	Dalapon	3.015	2.517	1418.5E6	469.3E6	703.104	673.708
2) T	3,5-DICHL...	6.403	5.957	1457.8E6	559.5E6	692.447	708.601
3) T	4-Nitroph...	7.013	6.387	543.2E6	216.2E6	679.751	703.583
5) T	DICAMBA	7.374	6.928	3998.8E6	1550.2E6	713.816	711.371
6) T	MCPD	7.544	7.024	297.8E6	123.3E6	70.383	71.382
7) T	MCPA	7.692	7.237	432.5E6	184.8E6	70.202	67.809
8) T	DICHLORPROP	8.066	7.548	942.9E6	383.4E6	700.642	707.606
9) T	2,4-D	8.296	7.839	1060.1E6	447.5E6	701.034	706.383
10) T	Pentachlo...	8.604	8.265	13998.9E6	5342.1E6	713.171	721.115
11) T	2,4,5-TP ...	9.153	8.626	5105.1E6	2103.2E6	706.027	720.797
12) T	2,4,5-T	9.438	9.003	5380.1E6	2090.4E6	705.782	721.651
13) T	2,4-DB	9.987	9.509	1017.8E6	266.6E6	700.246	704.821
14) T	DINOSEB	11.143	9.835	3747.9E6	1507.6E6	710.748	740.013
15) T	Picloram	10.965	10.813	6848.7E6	2856.9E6	686.185	709.499
16) T	DCPA	11.411	10.764	7447.3E6	3275.1E6	704.560	728.593

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS081621\
Data File : PS016428.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Aug 2021 00:33
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 17 06:32:09 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080221.M
Quant Title : 8080.M
QLast Update : Mon Aug 02 14:10:19 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

