

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS081622\
 Data File : PS020300.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Aug 2022 13:33
 Operator : AJ\MA
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 16 23:50:50 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080322.M
 Quant Title : 8080.M
 QLast Update : Thu Aug 04 02:05:37 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.181	6.757	1346.0E6	465.4E6	456.483	571.123 #
Target Compounds							
1) T	Dalapon	3.042	0.000	438317	0	<MDL	N.D. #
7) T	MCPA	7.639f	0.000	6058988	0	<MDL	N.D. #
8) T	DICHLORPROP	0.000	7.594f	0	2115073	N.D.	2.262
9) T	2,4-D	8.366f	0.000	15490668	0	5.144	N.D. #
11) T	2,4,5-TP ...	9.208f	8.580	20945019	2048005	1.229	<MDL #
12) T	2,4,5-T	9.408	0.000	17028535	0	<MDL	N.D. #
14) T	DINOSEB	11.104	0.000	15451336	0	<MDL	N.D. #
15) T	Picloram	10.947	10.846	13717803	4498741	<MDL	<MDL #
16) T	DCPA	11.337f	0.000	2272947	0	<MDL	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS081622\
Data File : PS020300.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Aug 2022 13:33
Operator : AJ\MA
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 16 23:50:50 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080322.M
Quant Title : 8080.M
QLast Update : Thu Aug 04 02:05:37 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA


R.T.: 7.181 min
Delta R.T.: -0.004 min
Response: 1346042362
Conc: 456.48 ng/ml

Instrument :
ECD_S
ClientSampleId :
I.BLK




#4 2,4-DCAA

R.T.: 6.757 min
Delta R.T.: -0.006 min
Response: 465439831
Conc: 571.12 ng/ml




#8 DICHLORPROP

R.T.: 0.000 min
Exp R.T. : 8.062 min
Response: 0
Conc: N.D.



#8 DICHLORPROP


R.T.: 7.594 min
Delta R.T.: 0.050 min
Response: 2115073
Conc: 2.26 ng/ml



#9 2,4-D


R.T.: 8.366 min
Delta R.T.: 0.067 min
Response: 15490668
Conc: 5.14 ng/ml

Instrument :
ECD_S
ClientSampleId :
I.BLK




#9 2,4-D

R.T.: 0.000 min
Exp R.T. : 7.840 min
Response: 0
Conc: N.D.



#11 2,4,5-TP (SILVEX)

R.T.: 9.208 min
Delta R.T.: 0.061 min
Response: 20945019
Conc: 1.23 ng/ml



#11 2,4,5-TP (SILVEX)

R.T.: 8.580 min
Delta R.T.: -0.040 min
Response: 2048005
Conc: N.D.