

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS081820\
 Data File : PS011822.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Aug 2020 10:14
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 18 13:01:37 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081120.M
 Quant Title : 8080.M
 QLast Update : Tue Aug 11 15:43:16 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.219	6.915	814.2E6	438.5E6	746.485	820.041
Target Compounds							
1) T	Dalapon	3.036	2.624	905.9E6	570.1E6	727.258	754.188
2) T	3,5-DICHL...	6.433	6.067	1037.8E6	619.3E6	715.882	752.866
3) T	4-Nitroph...	7.050	6.523	359.9E6	249.8E6	611.048	685.199
5) T	DICAMBA	7.409	7.083	3077.5E6	1691.4E6	751.337	782.888
6) T	MCPD	7.579	7.181	240.8E6	140.2E6	75.380	79.330
7) T	MCPA	7.729	7.396	336.3E6	209.6E6	73.942	76.477
8) T	DICHLORPROP	8.101	7.709	736.1E6	440.7E6	732.709	780.558
9) T	2,4-D	8.331	8.003	800.6E6	572.1E6	709.038	778.735
10) T	Pentachlo...	8.641	8.432	10550.1E6	5041.2E6	733.588	741.923
11) T	2,4,5-TP ...	9.190	8.793	4258.2E6	2281.5E6	732.306	780.375
12) T	2,4,5-T	9.476	9.172	4285.5E6	2283.6E6	723.185	771.476
13) T	2,4-DB	10.029	9.678	541.2E6	320.7E6	681.070	766.814
14) T	DINOSEB	11.187	10.011	3358.1E6	1671.5E6	715.460	786.226
15) T	Picloram	11.012	10.997	5146.0E6	3330.9E6	658.424	734.896
16) T	DCPA	11.453	10.939	5275.0E6	2862.4E6	693.761	778.485

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS081820\
Data File : PS011822.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Aug 2020 10:14
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 13:01:37 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081120.M
Quant Title : 8080.M
QLast Update : Tue Aug 11 15:43:16 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

