

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS081821\
 Data File : PS016481.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Aug 2021 15:33
 Operator : DD\AJ
 Sample : PB138518BSD
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB138518BSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 19 00:50:31 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080221.M
 Quant Title : 8080.M
 QLast Update : Mon Aug 02 14:10:19 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.184	6.765	761.0E6	268.9E6	517.330	515.437
Target Compounds							
1) T	Dalapon	3.014	2.517	901.1E6	299.3E6	446.633	429.657
2) T	3,5-DICHL...	6.403	5.957	1016.2E6	386.8E6	482.718	489.844
3) T	4-Nitroph...	7.015	6.387	352.7E6	146.9E6	441.288	478.169
5) T	DICAMBA	7.374	6.927	2787.4E6	1089.8E6	497.571	500.122
6) T	MCPD	7.541	7.022	199.0E6	84579219	47.013	48.956
7) T	MCPA	7.689	7.233	295.1E6	129.9E6	47.905	47.668
8) T	DICHLORPROP	8.066	7.548	659.3E6	268.5E6	489.921	495.625
9) T	2,4-D	8.296	7.839	727.2E6	307.2E6	480.915	484.812
10) T	Pentachlo...	8.603	8.265	10118.2E6	3930.9E6	515.466	530.616
11) T	2,4,5-TP ...	9.153	8.625	3611.3E6	1492.5E6	499.429	511.518
12) T	2,4,5-T	9.438	9.003	3736.6E6	1453.4E6	490.184	501.754
13) T	2,4-DB	9.990	9.509	642.7E6	179.1E6	442.205	473.460
14) T	DINOSEB	11.143	9.834	2657.5E6	1054.3E6	503.966	517.513
15) T	Picloram	10.966	10.814	4242.4E6	1758.8E6	425.051	436.790
16) T	DCPA	11.411	10.764	5362.2E6	2356.8E6	507.296	524.295

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS081821\
Data File : PS016481.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Aug 2021 15:33
Operator : DD\AJ
Sample : PB138518BSD
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PB138518BSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 19 00:50:31 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080221.M
Quant Title : 8080.M
QLast Update : Mon Aug 02 14:10:19 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

