

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS081924\
 Data File : PS027487.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Aug 2024 00:41
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 08/20/2024
 Supervised By : Ankita Jodhani 08/20/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 20 03:38:55 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081924.M
 Quant Title : 8080.M
 QLast Update : Mon Aug 19 14:50:32 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.164	7.738	1860.1E6	1407.5E6	719.231	712.950
Target Compounds							
1) T	Dalapon	2.592	2.691	2132.0E6	2322.9E6	642.059	635.877
2) T	3,5-DICHL...	6.345	6.694	2728.8E6	1935.2E6	665.743	667.968
3) T	4-Nitroph...	6.961	7.267	1014.1E6	872.5E6	651.236	653.822
5) T	DICAMBA	7.345	7.936	7680.3E6	6232.7E6	677.137	681.303
6) T	MCPD	7.524	8.038	528.3E6	420.1E6	68.942	69.010
7) T	MCPA	7.669	8.281	684.1E6	567.0E6	69.540	69.286m
8) T	DICHLORPROP	8.037	8.648	1852.0E6	1512.2E6	674.649	680.072
9) T	2,4-D	8.262	8.977	1988.2E6	1687.2E6	682.145	681.023
10) T	Pentachlo...	8.552	9.504	28912.7E6	23509.1E6	701.760	689.467
11) T	2,4,5-TP ...	9.123	9.877	10963.0E6	9350.8E6	695.095	684.826
12) T	2,4,5-T	9.410	10.296	10618.8E6	9231.6E6	698.324	683.369
13) T	2,4-DB	9.978	10.859	1494.2E6	1154.3E6	698.849	683.245
14) T	DINOSEB	11.164	11.238	8187.1E6	6205.7E6	682.569	664.533
15) T	Picloram	10.980	12.329	14051.2E6	13391.7E6	699.343	684.043
16) T	DCPA	11.461	12.278	13389.0E6	10169.4E6	713.835	690.673

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS081924\
Data File : PS027487.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Aug 2024 00:41
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :

ECD_S

ClientSampleId :

HSTDCCC750

Manual Integrations

APPROVED

Reviewed By : Abdul Mirza 08/20/2024
Supervised By : Ankita Jodhani 08/20/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 20 03:38:55 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081924.M
Quant Title : 8080.M
QLast Update : Mon Aug 19 14:50:32 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

