

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS082021\
 Data File : PS016534.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Aug 2021 16:26
 Operator : DD\AJ
 Sample : M3426-13MS
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 SWCS-10-0510MS

Manual Integrations
 APPROVED

mohammad
 8/23/2021 2:16:48 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 20 23:14:37 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080221.M
 Quant Title : 8080.M
 QLast Update : Mon Aug 02 14:10:19 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.182	6.764	541.1E6	198.0E6	367.787	379.632
Target Compounds							
1) T	Dalapon	3.011	2.515	544.6E6	195.6E6	269.914	280.812m
2) T	3,5-DICHL...	6.401	5.956	546.6E6	193.3E6	259.620	244.784
3) T	4-Nitroph...	7.011	6.385	197.7E6	77869186	247.389	253.408m
5) T	DICAMBA	7.372	6.926	1525.1E6	588.7E6	272.234	270.162
6) T	MCPD	7.539	7.018	96925873	43274690	22.904	25.048
7) T	MCPA	7.686	7.229	170.5E6	81629069	27.673	29.945
8) T	DICHLORPROP	8.064	7.546	355.8E6	138.2E6	264.372m	255.136
9) T	2,4-D	8.293	7.836	480.3E6	199.3E6	317.647	314.567m
10) T	Pentachlo...	8.602	8.264	3817.4E6	1489.2E6	194.474	201.020
11) T	2,4,5-TP ...	9.151	8.624	1757.0E6	742.5E6	242.985	254.459
12) T	2,4,5-T	9.435	9.002	1839.3E6	762.5E6	241.280	263.224
13) T	2,4-DB	9.987	9.507	188.0E6	84077071	129.331	222.249 #
14) T	DINOSEB	11.142	9.832	115.7E6	59976716	21.950	29.439 #
15) T	Picloram	10.963	10.811	2976.2E6	1252.9E6	298.187	311.147
16) T	DCPA	11.411	10.763	2344.8E6	1233.1E6	221.837	274.327

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS082021\
Data File : PS016534.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Aug 2021 16:26
Operator : DD\AJ
Sample : M3426-13MS
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
SWCS-10-0510MS

Manual Integrations
APPROVED

mohammad
8/23/2021 2:16:48 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 20 23:14:37 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080221.M
Quant Title : 8080.M
QLast Update : Mon Aug 02 14:10:19 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

