

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS082120\
 Data File : PS011878.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Aug 2020 19:04
 Operator : DD\AJ
 Sample : L3506-10MSD
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 OR-03-082020MSD

Manual Integrations
 APPROVED

Ankita
 8/24/2020 10:19:59 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 24 00:50:16 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081120.M
 Quant Title : 8080.M
 QLast Update : Tue Aug 11 15:43:16 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.218	6.915	244.7E6	156.8E6	224.314	293.108 #
Target Compounds							
1) T	Dalapon	3.033	2.625	511.9E6	286.9E6	410.911	379.527
2) T	3,5-DICHL...	6.432	6.066	615.9E6	349.2E6	424.855	424.477m
5) T	DICAMBA	7.408	7.082	1910.0E6	1092.3E6	466.316	505.570
6) T	MCP	7.576	7.177	216.8E6	85281496	67.850	48.252 #
7) T	MCPA	7.724	7.391	297.7E6	161.8E6	65.448	59.065m
8) T	DICHLORPROP	8.099	7.708	495.6E6	331.3E6	493.286	586.728
9) T	2,4-D	8.329	8.002	616.7E6	420.6E6	546.213	572.544
10) T	Pentachlo...	8.639	8.431	4283.2E6	2315.0E6	297.828	340.706
11) T	2,4,5-TP ...	9.189	8.787	3025.6E6	3120.9E6	520.324	1067.480 #
12) T	2,4,5-T	9.474	9.172	2917.7E6	1625.2E6	492.362	549.044
13) T	2,4-DB	10.027	9.678	311.6E6	220.0E6	392.174	525.925 #
14) T	DINOSEB	11.185	10.010	1023.0E6	522.9E6	217.953	245.945
15) T	Picloram	11.011	10.996	2295.8E6	1585.5E6	293.746	349.809
16) T	DCPA	11.451	10.938	2499.9E6	1456.6E6	328.790	396.142

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS081120\
Data File : PS011878.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Aug 2020 19:04
Operator : DD\AJ
Sample : L3506-10MSD
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
OR-03-082020MSD

Manual Integrations
APPROVED

Ankita
8/24/2020 10:19:59 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 24 00:50:16 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081120.M
Quant Title : 8080.M
QLast Update : Tue Aug 11 15:43:16 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

