

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\
 Method File : PS082218.M
 Title : 8080.M
 Last Update : Thu Aug 23 02:55:24 2018
 Response Via : Initial Calibration

Calibration Files

750 =PS001594.D 1000 =PS001595.D 1500 =PS001596.D
 500 =PS001593.D 200 =PS001592.D

	Compound	750	1000	1500	500	200	Avg		%RSD

1) T	Dalapon	2.575	2.402	2.339	2.549	2.703	2.514	E5	5.75
2) T	3,5-DICHLOROBENZOI...	2.734	2.568	2.482	2.684	2.924	2.678	E5	6.31
3) T	4-Nitrophenol	1.258	1.181	1.132	1.256	1.338	1.233	E5	6.42
4) S	2,4-DCAA	1.878	1.761	1.691	1.844	2.055	1.846	E5	7.47
5) T	DICAMBA	7.186	6.825	6.688	6.933	7.293	6.985	E5	3.59
6) T	MCPP	5.336	5.136	5.033	5.108	4.545	5.032	E5	5.85
7) T	MCPA	8.656	8.215	7.939	8.346	8.589	8.349	E5	3.48
8) T	DICHLORPROP	1.907	1.784	1.702	1.924	2.126	1.888	E5	8.53
9) T	2,4-D	2.406	2.267	2.183	2.389	2.611	2.371	E5	6.84
10) T	Pentachlorophenol	2.595	2.444	2.308	2.526	2.595	2.494	E6	4.85
11) T	2,4,5-TP (SILVEX)	9.607	9.089	8.753	9.349	9.862	9.332	E5	4.64
12) T	2,4,5-T	9.768	9.260	8.944	9.370	9.975	9.463	E5	4.34
13) T	2,4-DB	1.366	1.267	1.238	1.360	1.484	1.343	E5	7.22
14) T	DINOSEB	6.756	6.355	6.139	6.452	6.925	6.525	E5	4.83
15) T	Picloram	1.460	1.406	1.362	1.374	1.401	1.401	E6	2.70
16) T	DCPA	1.181	1.116	1.075	1.123	1.179	1.135	E6	3.99

Signal #2 Calibration Files

750 =PS001594.D 1000 =PS001595.D 1500 =PS001596.D
 500 =PS001593.D 200 =PS001592.D

	Compound	750	1000	1500	500	200	Avg		%RSD

1) T	Dalapon	2.452	2.329	2.283	2.425	2.478	2.393	E5	3.49
2) T	3,5-DICHLOROBENZOI...	2.615	2.497	2.432	2.540	2.693	2.556	E5	3.98
3) T	4-Nitrophenol	1.088	1.034	1.000	1.054	1.095	1.054	E5	3.74
4) S	2,4-DCAA	2.068	1.977	1.938	2.017	2.126	2.025	E5	3.66
5) T	DICAMBA	7.289	7.044	6.965	6.891	6.878	7.013	E5	2.39
6) T	MCPP	4.986	4.899	4.947	4.539	3.653	4.605	E5	12.19
7) T	MCPA	7.749	7.578	7.612	7.213	6.289	7.288	E5	8.14
8) T	DICHLORPROP	1.915	1.832	1.785	1.848	1.933	1.863	E5	3.27
9) T	2,4-D	2.550	2.444	2.393	2.438	2.561	2.477	E5	3.00
10) T	Pentachlorophenol	2.821	2.692	2.586	2.718	2.703	2.704	E6	3.09
11) T	2,4,5-TP (SILVEX)	1.104	1.062	1.038	1.054	1.056	1.063	E6	2.33
12) T	2,4,5-T	8.960	8.697	8.472	8.586	8.478	8.639	E5	2.33
13) T	2,4-DB	1.370	1.318	1.306	1.303	1.334	1.326	E5	2.07
14) T	DINOSEB	7.650	7.329	7.167	7.336	7.496	7.396	E5	2.49
15) T	Picloram	1.358	1.311	1.276	1.293	1.283	1.304	E6	2.51
16) T	DCPA	1.497	1.477	1.459	1.377	1.329	1.428	E6	5.02

(#) = Out of Range