

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS082224\
 Data File : PS027503.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Aug 2024 12:43
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 23 07:42:19 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081924.M
 Quant Title : 8080.M
 QLast Update : Mon Aug 19 14:50:32 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.159	7.735	1951.6E6	1487.8E6	754.587	753.615
Target Compounds							
1) T	Dalapon	2.593	2.693	2119.8E6	2228.7E6	638.387	610.095
2) T	3,5-DICHL...	6.342	6.692	2838.4E6	2055.1E6	692.478	709.350
3) T	4-Nitroph...	6.956	7.265	1083.9E6	917.1E6	696.039	687.305
5) T	DICAMBA	7.340	7.933	7820.4E6	6600.5E6	689.486	721.504
6) T	MCPD	7.519	8.035	550.6E6	449.3E6	71.850	73.815
7) T	MCPA	7.665	8.279	723.7E6	567.3E6	73.563	69.321
8) T	DICHLORPROP	8.032	8.645	1939.6E6	1577.5E6	706.549	709.429
9) T	2,4-D	8.257	8.974	2083.2E6	1780.0E6	714.747	718.445
10) T	Pentachlo...	8.546	9.500	30503.2E6	24929.7E6	740.365	731.129
11) T	2,4,5-TP ...	9.117	9.873	11567.6E6	9897.9E6	733.426	724.895
12) T	2,4,5-T	9.404	10.292	11204.8E6	9787.0E6	736.859	724.482
13) T	2,4-DB	9.971	10.855	1554.9E6	1215.0E6	727.208	719.176
14) T	DINOSEB	11.157	11.234	8887.8E6	6775.3E6	740.989	725.529
15) T	Picloram	10.975	12.325	14826.6E6	14331.6E6	737.938	732.055
16) T	DCPA	11.454	12.273	14059.3E6	10909.4E6	749.569	740.931

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS082224\
Data File : PS027503.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Aug 2024 12:43
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 23 07:42:19 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081924.M
Quant Title : 8080.M
QLast Update : Mon Aug 19 14:50:32 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

