

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS082419\
 Data File : PS006331.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Aug 2019 23:23
 Operator : SM\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 26 00:53:54 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081419.M
 Quant Title : 8080.M
 QLast Update : Wed Aug 14 15:09:52 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.493	5.222	638.1E6	349.0E6	739.259	709.599
Target Compounds							
1) T	Dalapon	2.058	1.768	683.2E6	381.9E6	627.031	608.627
2) T	3,5-DICHL...	4.860	4.525	729.3E6	403.7E6	676.489	666.453
3) T	4-Nitroph...	5.354	4.909	246.3E6	162.5E6	656.088	655.942
5) T	DICAMBA	5.643	5.349	2199.4E6	1230.8E6	684.831	667.264
6) T	MCPD	5.793	5.432	149.9E6	86988837	66.621	70.317
7) T	MCPA	5.926	5.610	221.9E6	147.6E6	70.274	72.243
8) T	DICHLORPROP	6.275	5.891	524.1E6	300.8E6	691.147	676.466
9) T	2,4-D	6.495	6.164	493.2E6	370.5E6	605.261	674.074
10) T	Pentachlo...	6.748	6.527	8455.0E6	4709.4E6	715.258	696.662
11) T	2,4,5-TP ...	7.285	6.892	3217.0E6	1805.6E6	705.211	693.889
12) T	2,4,5-T	7.561	7.252	3657.3E6	1614.4E6	739.661	649.363
13) T	2,4-DB	8.092	7.739	441.7E6	222.2E6	709.194	667.812
14) T	DINOSEB	9.175	8.047	2776.8E6	1319.9E6	758.530	687.134
15) T	Picloram	9.015	8.986	4765.6E6	2820.3E6	640.486	625.231
16) T	DCPA	9.448	8.942	4237.9E6	2477.8E6	731.228	710.548

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS082419\
Data File : PS006331.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Aug 2019 23:23
Operator : SM\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 26 00:53:54 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081419.M
Quant Title : 8080.M
QLast Update : Wed Aug 14 15:09:52 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

