

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS082423\
 Data File : PS024150.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Aug 2023 09:38
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 25 05:01:24 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080823.M
 Quant Title : 8080.M
 QLast Update : Tue Aug 08 13:26:17 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.365	6.794	1433.6E6	484.1E6	723.828	719.738
Target Compounds							
1) T	Dalapon	2.480	2.546	1380.0E6	670.5E6	628.911	613.232
2) T	3,5-DICHL...	5.738	5.977	2215.3E6	720.3E6	730.835m	687.465
3) T	4-Nitroph...	6.202	6.406	821.9E6	277.3E6	742.597	700.544
5) T	DICAMBA	6.507	6.957	6074.2E6	1997.6E6	690.494	681.124
6) T	MCPD	6.660	7.053	426.0E6	126.0E6	68.315	65.119m
7) T	MCPA	6.782	7.261	593.2E6	183.6E6	70.615	62.766
8) T	DICHLORPROP	7.096	7.573	1534.6E6	509.8E6	682.016	697.258
9) T	2,4-D	7.292	7.857	1810.7E6	602.7E6	730.704	707.480
10) T	Pentachlo...	7.523	8.283	22387.9E6	7547.5E6	714.173	702.431
11) T	2,4,5-TP ...	8.044	8.642	9387.9E6	2923.2E6	717.872	710.977
12) T	2,4,5-T	8.299	9.012	8564.9E6	2812.9E6	764.445	744.390
13) T	2,4-DB	8.804	9.516	1381.9E6	349.6E6	776.502	766.930
14) T	DINOSEB	9.859	9.844	6710.7E6	2109.8E6	723.935	703.489
15) T	Picloram	9.683	10.798	11035.5E6	3955.7E6	853.518	793.864
16) T	DCPA	10.124	10.768	11266.3E6	3714.0E6	716.135	722.424

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS082423\
Data File : PS024150.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Aug 2023 09:38
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 25 05:01:24 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080823.M
Quant Title : 8080.M
QLast Update : Tue Aug 08 13:26:17 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

