

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS082718\
 Data File : PS001625.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Aug 2018 00:06
 Operator : UA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Sohil
 8/28/2018 9:42:54 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 28 00:56:06 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS082218.M
 Quant Title : 8080.M
 QLast Update : Thu Aug 23 02:55:24 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.483	10.982	144.8E6	160.8E6	784.386	794.052
Target Compounds							
1) T	Dalapon	2.949	2.410	181.6E6	170.9E6	722.516	713.906
2) T	3,5-DICHL...	9.905	9.130	196.1E6	186.4E6	732.338	729.239m
3) T	4-Nitroph...	11.073	10.127	89116628	76587913	722.807	726.617
5) T	DICAMBA	11.796	11.268	519.9E6	533.1E6	744.300	760.105
6) T	MCPD	12.137	11.495	39532421	36448784	78.568	79.154
7) T	MCPA	12.387	11.871	63774534	56266627	76.386	77.202
8) T	DICHLORPROP	12.985	12.393	140.4E6	140.1E6	743.495	752.359
9) T	2,4-D	13.342	12.864	177.5E6	187.8E6	748.651	757.960
10) T	Pentachlo...	13.725	13.415	1918.1E6	2095.4E6	769.181	774.862
11) T	2,4,5-TP ...	14.596	14.032	717.1E6	826.3E6	768.409	777.478
12) T	2,4,5-T	14.992	14.600	732.7E6	673.1E6	774.305	779.162
13) T	2,4-DB	15.741	15.249	101.5E6	103.5E6	755.655	780.852
14) T	DINOSEB	17.233	15.679	499.0E6	566.0E6	764.727	765.275
15) T	Picloram	16.992	16.861	1089.7E6	1024.7E6	778.028	785.868
16) T	DCPA	17.573	16.915	894.1E6	1133.8E6	787.965	794.102

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS082718\
 Data File : PS001625.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Aug 2018 00:06
 Operator : UA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Sohil
 8/28/2018 9:42:54 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 28 00:56:06 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS082218.M
 Quant Title : 8080.M
 QLast Update : Thu Aug 23 02:55:24 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

