

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS082819\
 Data File : PS006368.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Aug 2019 14:16
 Operator : SM\AJ
 Sample : K4086-08MS
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 NB-07-082619MS

Manual Integrations
 APPROVED

Ankita
 8/29/2019 2:03:55 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 28 15:17:15 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081419.M
 Quant Title : 8080.M
 QLast Update : Wed Aug 14 15:09:52 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.525	5.254	687.9E6	392.1E6	796.971	797.206m
Target Compounds							
1) T	Dalapon	2.086	1.797	572.2E6	477.6E6	525.143	761.178 #
2) T	3,5-DICHL...	4.891	4.557	584.9E6	312.2E6	542.479m	515.401m
3) T	4-Nitroph...	5.385	4.941	188.2E6	138.9E6	501.429	560.628m
5) T	DICAMBA	5.677	5.383	1837.8E6	1091.3E6	572.213	591.673
6) T	MCPD	5.825	5.465	137.8E6	69098371	61.221m	55.855
7) T	MCPA	5.959	5.643	148.6E6	106.0E6	47.055	51.851
8) T	DICHLORPROP	6.308	5.924	554.6E6	281.8E6	731.348	633.728
9) T	2,4-D	6.526	6.196	582.4E6	389.7E6	714.807	708.924
10) T	Pentachlo...	6.782	6.561	7674.7E6	4321.4E6	649.249	639.264
11) T	2,4,5-TP ...	7.318	6.925	3095.7E6	1725.1E6	678.610	662.976
12) T	2,4,5-T	7.593	7.285	3625.6E6	1655.0E6	733.245	665.697
13) T	2,4-DB	8.124	7.771	508.6E6	199.0E6	816.681	597.866 #
14) T	DINOSEB	9.209	8.080	2639.8E6	1278.5E6	721.108	665.597
15) T	Picloram	9.048	9.019	4429.7E6	2595.3E6	595.337	575.342
16) T	DCPA	9.480	8.975	4123.1E6	2367.0E6	711.423	678.785

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS082819\
Data File : PS006368.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Aug 2019 14:16
Operator : SM\AJ
Sample : K4086-08MS
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
NB-07-082619MS

Manual Integrations
APPROVED

Ankita
8/29/2019 2:03:55 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 28 15:17:15 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081419.M
Quant Title : 8080.M
QLast Update : Wed Aug 14 15:09:52 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

