

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS082823\
 Data File : PS024177.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Aug 2023 17:54
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 29 03:25:58 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS082823.M
 Quant Title : 8080.M
 QLast Update : Mon Aug 28 16:48:10 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.364	6.793	1466.2E6	499.7E6	735.913	743.987
Target Compounds							
1) T	Dalapon	2.481	2.549	1462.8E6	708.5E6	673.013	671.919
2) T	3,5-DICHL...	5.738	5.975	2095.0E6	726.9E6	689.209	697.184
3) T	4-Nitroph...	6.201	6.405	839.7E6	284.7E6	675.875	675.061
5) T	DICAMBA	6.506	6.956	6160.1E6	2051.6E6	706.659	712.312
6) T	MCPD	6.658	7.052	453.8E6	138.7E6	72.880	72.379
7) T	MCPA	6.781	7.260	616.2E6	207.5E6	70.999	70.988
8) T	DICHLORPROP	7.095	7.571	1579.2E6	508.2E6	694.672	699.453
9) T	2,4-D	7.291	7.855	1845.4E6	626.2E6	693.797	702.871
10) T	Pentachlo...	7.521	8.281	22628.2E6	7614.1E6	723.580	718.671
11) T	2,4,5-TP ...	8.043	8.640	9509.9E6	2971.8E6	711.455	716.232
12) T	2,4,5-T	8.298	9.010	8881.6E6	2886.5E6	716.831	719.274
13) T	2,4-DB	8.803	9.514	1538.1E6	345.1E6	713.539	750.803
14) T	DINOSEB	9.858	9.842	6787.1E6	2114.6E6	702.418	707.390
15) T	Picloram	9.681	10.796	11807.5E6	4327.1E6	723.001	728.617
16) T	DCPA	10.123	10.765	11585.1E6	3725.1E6	715.399	726.153

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS082823\
Data File : PS024177.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Aug 2023 17:54
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 29 03:25:58 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS082823.M
Quant Title : 8080.M
QLast Update : Mon Aug 28 16:48:10 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

