

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS082824\
 Data File : PS027537.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Aug 2024 21:12
 Operator : AR\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 29 05:08:13 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081924.M
 Quant Title : 8080.M
 QLast Update : Mon Aug 19 14:50:32 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.155	7.736	1309.6E6	1011.8E6	506.376	512.524
Target Compounds							
1) T	Dalapon	0.000	2.711	0	180080	N.D.	<MDL
2) T	3,5-DICHL...	0.000	6.683	0	419538	N.D.	<MDL
3) T	4-Nitroph...	6.950	0.000	9374539	0	6.020	N.D. #
5) T	DICAMBA	0.000	7.929	0	7423586	N.D.	<MDL
6) T	MCP	7.570f	8.037	24712536	2577327	3.225	<MDL #
8) T	DICHLORPROP	8.050	8.696f	1052383	9842305	<MDL	4.426 #
9) T	2,4-D	8.269	9.015f	31232037	170794	10.716	<MDL #
10) T	Pentachlo...	0.000	9.460f	0	2046453	N.D.	<MDL
11) T	2,4,5-TP ...	9.086f	9.878	92679402	136713	5.876	<MDL #
12) T	2,4,5-T	0.000	10.296	0	1273031	N.D.	<MDL
13) T	2,4-DB	0.000	10.854	0	94072	N.D.	<MDL
14) T	DINOSEB	0.000	11.243	0	827834	N.D.	<MDL
15) T	Picloram	10.952f	0.000	20060560	0	<MDL	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS082824\
Data File : PS027537.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Aug 2024 21:12
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 29 05:08:13 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081924.M
Quant Title : 8080.M
QLast Update : Mon Aug 19 14:50:32 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#3 4-Nitrophenol


R.T.: 6.950 min
Delta R.T.: -0.014 min
Response: 9374539
Conc: 6.02 ng/ml

Instrument :
ECD_S
ClientSampleId :
I.BLK




#3 4-Nitrophenol

R.T.: 0.000 min
Exp R.T. : 7.267 min
Response: 0
Conc: N.D.




#4 2,4-DCAA

R.T.: 7.155 min
Delta R.T.: -0.011 min
Response: 1309620605
Conc: 506.38 ng/ml



#4 2,4-DCAA


R.T.: 7.736 min
Delta R.T.: -0.002 min
Response: 1011842994
Conc: 512.52 ng/ml



#6 MCP


R.T.: 7.570 min
Delta R.T.: 0.048 min
Response: 24712536
Conc: 3.23 ug/ml

Instrument :
ECD_S
ClientSampleId :
I.BLK




#6 MCP

R.T.: 8.037 min
Delta R.T.: 0.003 min
Response: 2577327
Conc: N.D.




#8 DICHLORPROP

R.T.: 8.050 min
Delta R.T.: 0.011 min
Response: 1052383
Conc: N.D.



#8 DICHLORPROP


R.T.: 8.696 min
Delta R.T.: 0.048 min
Response: 9842305
Conc: 4.43 ng/ml



#9 2,4-D


R.T.: 8.269 min
Delta R.T.: 0.003 min
Response: 31232037
Conc: 10.72 ng/ml

Instrument :
ECD_S
ClientSampleId :
I.BLK




#9 2,4-D

R.T.: 9.015 min
Delta R.T.: 0.037 min
Response: 170794
Conc: N.D.



#11 2,4,5-TP (SILVEX)

R.T.: 9.086 min
Delta R.T.: -0.040 min
Response: 92679402
Conc: 5.88 ng/ml



#11 2,4,5-TP (SILVEX)

R.T.: 9.878 min
Delta R.T.: 0.000 min
Response: 136713
Conc: N.D.