

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS082824\
 Data File : PS027550.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Aug 2024 02:53
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 08/29/2024
 Supervised By :Ankita Jodhani 08/29/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 29 05:12:44 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081924.M
 Quant Title : 8080.M
 QLast Update : Mon Aug 19 14:50:32 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.154	7.735	1881.3E6	1429.7E6	727.413	724.179
Target Compounds							
1) T	Dalapon	2.589	2.693	2065.0E6	2289.1E6	621.887	626.616
2) T	3,5-DICHL...	6.338	6.692	2732.1E6	1956.4E6	666.549m	675.305
3) T	4-Nitroph...	6.952	7.264	1034.0E6	892.6E6	664.036	668.924m
5) T	DICAMBA	7.335	7.933	7683.1E6	6315.7E6	677.379m	690.367
6) T	MCP	7.514	8.035	547.9E6	427.3E6	71.505m	70.192
7) T	MCPA	7.659	8.277	690.0E6	570.7E6	70.136	69.741m
8) T	DICHLORPROP	8.026	8.644	1863.4E6	1523.2E6	678.788	685.011m
9) T	2,4-D	8.251	8.973	2006.4E6	1706.9E6	688.395	688.962
10) T	Pentachlo...	8.540	9.499	29008.5E6	23684.3E6	704.087	694.604
11) T	2,4,5-TP ...	9.110	9.873	11018.1E6	9413.7E6	698.588	689.434m
12) T	2,4,5-T	9.397	10.291	10802.6E6	9351.7E6	710.411	692.260
13) T	2,4-DB	9.963	10.855	1523.0E6	1160.7E6	712.308	687.067
14) T	DINOSEB	11.147	11.233	8300.3E6	6289.5E6	692.009	673.514m
15) T	Picloram	10.964	12.322	14221.8E6	13514.9E6	707.834	690.335
16) T	DCPA	11.446	12.273	13436.4E6	10420.9E6	716.359m	707.750

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS082824\
Data File : PS027550.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Aug 2024 02:53
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 08/29/2024
Supervised By : Ankita Jodhani 08/29/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 29 05:12:44 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081924.M
Quant Title : 8080.M
QLast Update : Mon Aug 19 14:50:32 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

