

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS083018\
 Data File : PS001646.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Aug 2018 11:44
 Operator : EI\SJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Sohil
 8/31/2018 10:45:11 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 31 00:59:39 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS082218.M
 Quant Title : 8080.M
 QLast Update : Thu Aug 23 02:55:24 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.472	10.962	130.8E6	146.0E6	708.805	720.692
Target Compounds							
1) T	Dalapon	2.935	2.387	162.9E6	148.5E6	647.862m	620.489m
2) T	3,5-DICHL...	9.893	9.108	176.7E6	167.9E6	659.912	657.182
3) T	4-Nitroph...	11.061	10.104	81275514	69138007	659.209	655.937m
5) T	DICAMBA	11.785	11.247	467.8E6	477.3E6	669.666	680.533m
6) T	MCPD	12.123	11.474	35738433	33420550	71.028m	72.578m
7) T	MCPA	12.375	11.851	54262395	51004224	64.993	69.982m
8) T	DICHLORPROP	12.975	12.374	127.6E6	124.1E6	675.889	666.441m
9) T	2,4-D	13.333	12.845	160.0E6	170.2E6	674.721	687.179m
10) T	Pentachlo...	13.714	13.398	1722.8E6	1909.4E6	690.865	706.070
11) T	2,4,5-TP ...	14.586	14.015	645.7E6	750.2E6	691.959	705.879
12) T	2,4,5-T	14.983	14.584	657.2E6	610.4E6	694.436	706.577
13) T	2,4-DB	15.732	15.233	93303910	93910073	694.622	708.165m
14) T	DINOSEB	17.223	15.663	450.1E6	514.9E6	689.862	696.175m
15) T	Picloram	16.983	16.846	986.8E6	935.9E6	704.575	717.704m
16) T	DCPA	17.564	16.901	807.0E6	1026.7E6	711.238m	719.076m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS083018\
Data File : PS001646.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Aug 2018 11:44
Operator : EI\SJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
HSTDCCC750

Manual Integrations
APPROVED

Sohil
8/31/2018 10:45:11 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 31 00:59:39 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS082218.M
Quant Title : 8080.M
QLast Update : Thu Aug 23 02:55:24 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

