

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS083024\
 Data File : PS027572.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Aug 2024 11:35
 Operator : AR\AJ
 Sample : PB163083BS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB163083BS

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 09/03/2024
 Supervised By :Ankita Jodhani 09/03/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 31 04:23:32 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081924.M
 Quant Title : 8080.M
 QLast Update : Mon Aug 19 14:50:32 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.151	7.734	1263.2E6	960.2E6	488.445	486.387
Target Compounds							
1) T	Dalapon	2.588	2.690	1358.1E6	1560.5E6	408.986	427.171
2) T	3,5-DICHL...	6.335	6.690	1818.5E6	1307.5E6	443.646m	451.321
3) T	4-Nitroph...	6.949	7.263	687.6E6	606.6E6	441.580	454.553m
5) T	DICAMBA	7.331	7.931	5055.2E6	4169.7E6	445.693m	455.792
6) T	MCPD	7.509	8.031	334.9E6	269.2E6	43.701m	44.228
7) T	MCPA	7.654	8.274	432.7E6	363.2E6	43.986	44.387
8) T	DICHLORPROP	8.021	8.642	1244.5E6	1027.3E6	453.348m	462.006m
9) T	2,4-D	8.247	8.971	1337.6E6	1146.6E6	458.918	462.799
10) T	Pentachlo...	8.536	9.498	19590.7E6	16259.6E6	475.499	476.857
11) T	2,4,5-TP ...	9.106	9.870	7357.5E6	6354.6E6	466.490	465.395m
12) T	2,4,5-T	9.393	10.289	7193.4E6	6272.0E6	473.057	464.286
13) T	2,4-DB	9.960	10.853	978.6E6	757.5E6	457.696	448.409
14) T	DINOSEB	11.143	11.231	5672.7E6	4320.9E6	472.938	462.701m
15) T	Picloram	10.962	12.322	8900.4E6	8473.3E6	442.982	432.813
16) T	DCPA	11.442	12.271	9027.3E6	7035.2E6	481.288m	477.804

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS083024\
Data File : PS027572.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Aug 2024 11:35
Operator : AR\AJ
Sample : PB163083BS
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PB163083BS

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 09/03/2024
Supervised By : Ankita Jodhani 09/03/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 31 04:23:32 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081924.M
Quant Title : 8080.M
QLast Update : Mon Aug 19 14:50:32 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

