

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS083121\
 Data File : PS016731.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Aug 2021 18:19
 Operator : DD\AJ
 Sample : HSTDICV750
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 ICVPS083121

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 01 02:22:20 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS083121.M
 Quant Title : 8080.M
 QLast Update : Wed Sep 01 02:20:15 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.182	6.763	1112.3E6	387.9E6	740.629	748.402
Target Compounds							
1) T	Dalapon	3.013	2.514	1543.1E6	466.1E6	684.839	693.353
2) T	3,5-DICHL...	6.401	5.956	1474.9E6	553.6E6	692.767	697.023
3) T	4-Nitroph...	7.012	6.386	527.8E6	204.3E6	682.618	690.120
5) T	DICAMBA	7.371	6.926	3998.3E6	1567.4E6	702.545	710.079
6) T	MCPD	7.540	7.022	302.3E6	121.7E6	71.676	71.419
7) T	MCPA	7.688	7.234	410.6E6	183.2E6	70.084	68.371
8) T	DICHLORPROP	8.063	7.545	952.0E6	382.7E6	694.143	703.097
9) T	2,4-D	8.294	7.837	1054.4E6	439.6E6	696.114	700.003
10) T	Pentachlo...	8.599	8.263	13764.6E6	5309.9E6	704.797	711.718
11) T	2,4,5-TP ...	9.151	8.624	5066.1E6	2103.4E6	705.567	715.513
12) T	2,4,5-T	9.435	9.002	5078.4E6	2012.9E6	709.719	716.306
13) T	2,4-DB	9.987	9.507	998.8E6	257.6E6	722.630	705.671
14) T	DINOSEB	11.140	9.832	3786.5E6	1495.5E6	694.437	713.224
15) T	Picloram	10.966	10.813	6543.6E6	2716.0E6	734.342	741.233
16) T	DCPA	11.407	10.762	7034.7E6	3090.0E6	706.217	712.216

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS083121\
Data File : PS016731.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Aug 2021 18:19
Operator : DD\AJ
Sample : HSTDICV750
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
ICVPS083121

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 01 02:22:20 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS083121.M
Quant Title : 8080.M
QLast Update : Wed Sep 01 02:20:15 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

