

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS090120\
 Data File : PS012024.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Sep 2020 21:40
 Operator : DD\AJ
 Sample : L3508-28
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 TR-08-082820

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 02 03:53:38 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081120.M
 Quant Title : 8080.M
 QLast Update : Tue Aug 11 15:43:16 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.217	6.914	284.0E6	176.5E6	260.322	330.044 #
Target Compounds							
1) T	Dalapon	2.998	2.586	41625133	89396664	33.416	118.273 #
2) T	3,5-DICHL...	0.000	6.095	0	88908725	N.D.	108.080
3) T	4-Nitroph...	7.055	0.000	3982415	0	6.761	N.D. #
5) T	DICAMBA	7.395	7.048	7874711	12234430	1.923	5.663 #
6) T	MCPD	7.570	0.000	7885108	0	2.468	N.D. #
7) T	MCPA	7.689	7.463	216.0E6	120.8E6	47.488	44.080
8) T	DICHLORPROP	0.000	7.736	0	14371292	N.D.	25.452
9) T	2,4-D	8.396	0.000	149.4E6	0	132.332	N.D. #
10) T	Pentachlo...	8.693	0.000	14888801	0	1.035	N.D. #
12) T	2,4,5-T	9.411	0.000	30658757	0	5.174	N.D. #
16) T	DCPA	11.506	0.000	11180827	0	1.470	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS090120\
Data File : PS012024.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Sep 2020 21:40
Operator : DD\AJ
Sample : L3508-28
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
TR-08-082820

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 02 03:53:38 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081120.M
Quant Title : 8080.M
QLast Update : Tue Aug 11 15:43:16 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 2.998 min
Delta R.T.: -0.037 min
Response: 41625133
Conc: 33.42 ng/ml

Instrument :
ECD_S
ClientSampleId :
TR-08-082820




#1 Dalapon

R.T.: 2.586 min
Delta R.T.: -0.035 min
Response: 89396664
Conc: 118.27 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 0.000 min
Exp R.T. : 6.434 min
Response: 0
Conc: N.D.



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 6.095 min
Delta R.T.: 0.029 min
Response: 88908725
Conc: 108.08 ng/ml



#3 4-Nitrophenol


R.T.: 7.055 min
Delta R.T.: 0.004 min
Response: 3982415
Conc: 6.76 ng/ml

Instrument :
ECD_S
ClientSampleId :
TR-08-082820




#3 4-Nitrophenol

R.T.: 0.000 min
Exp R.T. : 6.521 min
Response: 0
Conc: N.D.




#4 2,4-DCAA

R.T.: 7.217 min
Delta R.T.: -0.003 min
Response: 283951049
Conc: 260.32 ng/ml



#4 2,4-DCAA


R.T.: 6.914 min
Delta R.T.: 0.000 min
Response: 176503232
Conc: 330.04 ng/ml



#5 DICAMBA


R.T.: 7.395 min
Delta R.T.: -0.017 min
Response: 7874711
Conc: 1.92 ng/ml

Instrument :
ECD_S
ClientSampleId :
TR-08-082820




#5 DICAMBA

R.T.: 7.048 min
Delta R.T.: -0.036 min
Response: 12234430
Conc: 5.66 ng/ml




#6 MCPP

R.T.: 7.570 min
Delta R.T.: -0.016 min
Response: 7885108
Conc: 2.47 ug/ml



#6 MCPP


R.T.: 0.000 min
Exp R.T. : 7.188 min
Response: 0
Conc: N.D.



#7 MCPA


R.T.: 7.689 min
Delta R.T.: -0.049 min
Response: 215997681
Conc: 47.49 ug/ml

Instrument :
ECD_S
ClientSampleId :
TR-08-082820




#7 MCPA

R.T.: 7.463 min
Delta R.T.: 0.058 min
Response: 120786028
Conc: 44.08 ug/ml




#8 DICHLORPROP

R.T.: 0.000 min
Exp R.T. : 8.102 min
Response: 0
Conc: N.D.



#8 DICHLORPROP


R.T.: 7.736 min
Delta R.T.: 0.027 min
Response: 14371292
Conc: 25.45 ng/ml



#9 2,4-D


R.T.: 8.396 min
Delta R.T.: 0.064 min
Response: 149415237
Conc: 132.33 ng/ml

Instrument :
ECD_S
ClientSampleId :
TR-08-082820




#9 2,4-D

R.T.: 0.000 min
Exp R.T. : 8.003 min
Response: 0
Conc: N.D.




#10 Pentachlorophenol

R.T.: 8.693 min
Delta R.T.: 0.050 min
Response: 14888801
Conc: 1.04 ng/ml



#10 Pentachlorophenol


R.T.: 0.000 min
Exp R.T. : 8.434 min
Response: 0
Conc: N.D.



#12 2,4,5-T


R.T.: 9.411 min
Delta R.T.: -0.065 min
Response: 30658757
Conc: 5.17 ng/ml

Instrument :
ECD_S
ClientSampleId :
TR-08-082820




#12 2,4,5-T

R.T.: 0.000 min
Exp R.T. : 9.172 min
Response: 0
Conc: N.D.



#16 DCPA

R.T.: 11.506 min
Delta R.T.: 0.053 min
Response: 11180827
Conc: 1.47 ng/ml



#16 DCPA

R.T.: 0.000 min
Exp R.T. : 10.940 min
Response: 0
Conc: N.D.