

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS090121\
 Data File : PS016739.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Sep 2021 11:41
 Operator : DD\AJ
 Sample : M3550-01MS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 WC-1MS

Manual Integrations
 APPROVED

mohammad
 9/2/2021 2:48:00 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 02 02:59:53 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS083121.M
 Quant Title : 8080.M
 QLast Update : Wed Sep 01 10:13:14 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.181	6.763	352.9E6	101.1E6	234.994	195.121
Target Compounds							
1) T	Dalapon	3.008	2.513	1049.8E6	402.8E6	465.942m	599.115m#
2) T	3,5-DICHL...	6.399	5.955	901.4E6	317.8E6	423.399	400.150m
5) T	DICAMBA	7.370	6.924	2726.3E6	1048.8E6	479.030	475.132m
6) T	MCP	7.537	7.020	234.8E6	81495064	55.687	47.809
7) T	MCPA	7.685	7.230	364.0E6	141.1E6	62.138	52.668
8) T	DICHLORPROP	8.062	7.544	689.6E6	277.9E6	502.841	510.625
9) T	2,4-D	8.291	7.836	878.8E6	392.5E6	580.124	624.996
10) T	Pentachlo...	8.598	8.262	4839.2E6	1881.2E6	247.786	252.143
11) T	2,4,5-TP ...	9.148	8.618	3990.0E6	3514.9E6	555.701	1195.659 #
12) T	2,4,5-T	9.433	9.000	3797.9E6	1527.9E6	530.768	543.699
13) T	2,4-DB	9.985	9.506	479.7E6	202.6E6	347.063m	555.126 #
14) T	DINOSEB	11.138	9.832	1064.8E6	389.8E6	195.291m	185.904
15) T	Picloram	10.961	10.811	5621.3E6	1960.2E6	630.841m	534.951
16) T	DCPA	11.406	10.761	4775.9E6	2188.4E6	479.451	504.402

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS090121\
Data File : PS016739.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Sep 2021 11:41
Operator : DD\AJ
Sample : M3550-01MS
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
WC-1MS

Manual Integrations
APPROVED

mohammad
9/2/2021 2:48:00 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 02 02:59:53 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS083121.M
Quant Title : 8080.M
QLast Update : Wed Sep 01 10:13:14 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

