

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS090220\
 Data File : PS012061.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Sep 2020 22:43
 Operator : DD\AJ
 Sample : L3837-04
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 03 05:10:01 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081120.M
 Quant Title : 8080.M
 QLast Update : Tue Aug 11 15:43:16 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.219	6.915	213.1E6	132.0E6	195.406	246.906 #
Target Compounds							
1) T	Dalapon	2.997	2.585	35615812	61231028	28.592	81.010 #
2) T	3,5-DICHL...	0.000	6.096	0	73154733	N.D.	88.929
3) T	4-Nitroph...	6.984	0.000	7207976	0	12.237	N.D. #
5) T	DICAMBA	7.395	7.048	6721678	9075012	1.641	4.200 #
6) T	MCPD	7.571	0.000	8226741	0	2.575	N.D. #
7) T	MCPA	7.695	7.470	196.0E6	104.0E6	43.093	37.955
9) T	2,4-D	8.398	0.000	123.3E6	0	109.230	N.D. #
10) T	Pentachlo...	8.694	0.000	16094485	0	1.119	N.D. #
11) T	2,4,5-TP ...	0.000	8.788	0	165.0E6	N.D.	56.424
12) T	2,4,5-T	9.411	0.000	56435112	0	9.524	N.D. #
15) T	Picloram	11.008	0.000	17892819	0	2.289	N.D. #
16) T	DCPA	11.460	0.000	9242979	0	1.216	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS090220\
Data File : PS012061.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Sep 2020 22:43
Operator : DD\AJ
Sample : L3837-04
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 03 05:10:01 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081120.M
Quant Title : 8080.M
QLast Update : Tue Aug 11 15:43:16 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 2.997 min
Delta R.T.: -0.038 min
Response: 35615812
Conc: 28.59 ng/ml

Instrument :
ECD_S
ClientSampleId :




#1 Dalapon

R.T.: 2.585 min
Delta R.T.: -0.036 min
Response: 61231028
Conc: 81.01 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 0.000 min
Exp R.T. : 6.434 min
Response: 0
Conc: N.D.



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 6.096 min
Delta R.T.: 0.029 min
Response: 73154733
Conc: 88.93 ng/ml



#3 4-Nitrophenol


R.T.: 6.984 min
Delta R.T.: -0.067 min
Response: 7207976
Conc: 12.24 ng/ml

Instrument :
ECD_S
ClientSampleId :




#3 4-Nitrophenol

R.T.: 0.000 min
Exp R.T. : 6.521 min
Response: 0
Conc: N.D.




#4 2,4-DCAA

R.T.: 7.219 min
Delta R.T.: -0.001 min
Response: 213143208
Conc: 195.41 ng/ml



#4 2,4-DCAA


R.T.: 6.915 min
Delta R.T.: 0.000 min
Response: 132041783
Conc: 246.91 ng/ml



#5 DICAMBA


R.T.: 7.395 min
Delta R.T.: -0.016 min
Response: 6721678
Conc: 1.64 ng/ml

Instrument :
ECD_S
ClientSampleId :




#5 DICAMBA

R.T.: 7.048 min
Delta R.T.: -0.036 min
Response: 9075012
Conc: 4.20 ng/ml




#6 MCP

R.T.: 7.571 min
Delta R.T.: -0.015 min
Response: 8226741
Conc: 2.57 ug/ml



#6 MCP


R.T.: 0.000 min
Exp R.T.: 7.188 min
Response: 0
Conc: N.D.



#7 MCPA


R.T.: 7.695 min
Delta R.T.: -0.043 min
Response: 196009895
Conc: 43.09 ug/ml

Instrument :
ECD_S
ClientSampleId :




#7 MCPA

R.T.: 7.470 min
Delta R.T.: 0.065 min
Response: 104002436
Conc: 37.96 ug/ml




#9 2,4-D

R.T.: 8.398 min
Delta R.T.: 0.066 min
Response: 123330679
Conc: 109.23 ng/ml



#9 2,4-D


R.T.: 0.000 min
Exp R.T.: 8.003 min
Response: 0
Conc: N.D.



#10 Pentachlorophenol


R.T.: 8.694 min
Delta R.T.: 0.051 min
Response: 16094485
Conc: 1.12 ng/ml

Instrument :
ECD_S
ClientSampleId :




#10 Pentachlorophenol

R.T.: 0.000 min
Exp R.T. : 8.434 min
Response: 0
Conc: N.D.




#11 2,4,5-TP (SILVEX)

R.T.: 0.000 min
Exp R.T. : 9.191 min
Response: 0
Conc: N.D.



#11 2,4,5-TP (SILVEX)


R.T.: 8.788 min
Delta R.T.: -0.005 min
Response: 164965207
Conc: 56.42 ng/ml



#12 2,4,5-T


R.T.: 9.411 min
Delta R.T.: -0.064 min
Response: 56435112
Conc: 9.52 ng/ml

Instrument :
ECD_S
ClientSampleId :




#12 2,4,5-T

R.T.: 0.000 min
Exp R.T. : 9.172 min
Response: 0
Conc: N.D.




#15 Picloram

R.T.: 11.008 min
Delta R.T.: -0.001 min
Response: 17892819
Conc: 2.29 ng/ml



#15 Picloram


R.T.: 0.000 min
Exp R.T. : 10.995 min
Response: 0
Conc: N.D.



#16 DCPA

R.T.: 11.460 min
Delta R.T.: 0.006 min
Response: 9242979
Conc: 1.22 ng/ml

Instrument :
ECD_S
ClientSampleId :



#16 DCPA

R.T.: 0.000 min
Exp R.T. : 10.940 min
Response: 0
Conc: N.D.