

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS090222\
 Data File : PS020432.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Sep 2022 12:28
 Operator : AJ\MA
 Sample : HSTDICC1500
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC1500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 03 00:59:12 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS090222.M
 Quant Title : 8080.M
 QLast Update : Sat Sep 03 00:56:43 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.174	6.755	4973.3E6	1611.5E6	1477.282	1521.704
Target Compounds							
1) T	Dalapon	2.996	2.511	6605.5E6	1927.0E6	1365.000	1368.015
2) T	3,5-DICHL...	6.392	5.942	6092.5E6	2273.6E6	1356.277	1384.065
3) T	4-Nitroph...	7.004	6.376	2486.5E6	904.9E6	1460.842	1398.645
5) T	DICAMBA	7.363	6.918	17965.0E6	6393.8E6	1356.117	1380.404
6) T	MCPD	7.540	7.024	1607.7E6	539.6E6	144.581	144.295
7) T	MCPA	7.690	7.238	2184.4E6	758.4E6	141.456	141.454
8) T	DICHLORPROP	8.054	7.537	4833.9E6	1720.3E6	1386.127	1410.461
9) T	2,4-D	8.284	7.829	5809.1E6	2126.7E6	1330.538	1419.140
10) T	Pentachlo...	8.592	8.252	46393.5E6	19945.9E6	1162.418	1307.640
11) T	2,4,5-TP ...	9.138	8.612	26438.3E6	9424.9E6	1353.820	1373.019
12) T	2,4,5-T	9.424	8.989	28216.2E6	9148.6E6	1337.603	1390.348
13) T	2,4-DB	9.975	9.493	6430.1E6	1331.5E6	1449.927	1461.588
14) T	DINOSEB	11.124	9.817	27376.4E6	7013.1E6	1322.021	1382.977
15) T	Picloram	10.950	10.796	53396.5E6	13999.9E6	1357.110	1399.069
16) T	DCPA	11.394	10.744	39467.3E6	11411.4E6	1356.079	1375.139

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS090222\
Data File : PS020432.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Sep 2022 12:28
Operator : AJ\MA
Sample : HSTDICC1500
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC1500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 03 00:59:12 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS090222.M
Quant Title : 8080.M
QLast Update : Sat Sep 03 00:56:43 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

