

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS090225\
 Data File : PS031589.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Sep 2025 15:38
 Operator : AR\AJ
 Sample : PB169463BSD
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB169463BSD

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 09/03/2025
 Supervised By :mohammad ahmed 09/04/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 03 02:19:58 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080725.M
 Quant Title : 8080.M
 QLast Update : Fri Aug 08 03:17:54 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.287	7.744	1872.8E6	432.4E6	590.059	511.539
Target Compounds							
1) T	Dalapon	2.670	2.692	2513.7E6	1218.9E6	478.892	458.053
2) T	3,5-DICHL...	6.454	6.696	2568.5E6	719.1E6	544.362	531.914
3) T	4-Nitroph...	7.100	7.301	524.7E6	493.7E6	508.583	392.777
5) T	DICAMBA	7.470	7.938	7531.4E6	2929.3E6	539.252	515.474
6) T	MCPD	7.646	8.033	426.1E6	94125725	52.961m	53.244m
7) T	MCPA	7.796	8.280	499.6E6	126.9E6	50.150	45.239
8) T	DICHLORPROP	8.178	8.658	1716.0E6	657.2E6	540.908	492.288
9) T	2,4-D	8.416	9.003	1642.7E6	693.1E6	571.017m	485.589
10) T	Pentachlo...	8.709	9.508	28185.1E6	19327.0E6	587.038	554.919
11) T	2,4,5-TP ...	9.287	9.893	10391.4E6	7032.0E6	579.755	536.135
12) T	2,4,5-T	9.586	10.323	8616.6E6	5930.6E6	588.112	500.902
13) T	2,4-DB	10.161	10.893	840.1E6	397.1E6	437.144m	411.602
14) T	DINOSEB	11.357	11.266	7284.7E6	5045.6E6	574.574	533.537
15) T	Picloram	11.192	12.393	5690.8E6	7140.2E6	435.936m	407.321m
16) T	DCPA	11.655	12.306	13616.2E6	11360.5E6	597.744m	576.665m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS090225\
Data File : PS031589.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Sep 2025 15:38
Operator : AR\AJ
Sample : PB169463BSD
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PB169463BSD

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 09/03/2025
Supervised By : mohammad ahmed 09/04/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 03 02:19:58 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080725.M
Quant Title : 8080.M
QLast Update : Fri Aug 08 03:17:54 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

